



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93692

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/05/2018	09:10
2	J1789-01		1	55.0	131.0	03/05/2018	09:32
3	J1789-01DUP		1	55.0	131.0	03/05/2018	09:55

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB93692

WorkList Name : FP3518

WorkList ID : 109246

Date : 3/5/2018 8:10:30 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/09/2018	✓ Solid	J1789-01	✓ Flash Point	Cool 4 deg C	PSEG02	G31	19614	03/02/2018	1010A
03/07/2018	Water	J1792-12	✓ Flash Point	Cool 4 deg C	PSEG01	G32	TP-7	03/01/2018	1010A
03/07/2018	Water	J1792-14	✓ Flash Point	Cool 4 deg C	PSEG01	G32	TP-8	03/01/2018	1010A
03/07/2018	Water	J1792-16	✓ Flash Point	Cool 4 deg C	PSEG01	G32	TP-13	03/01/2018	1010A

Date/Time

Received by:

Relinquished by:

3/5/18 8:30 AM
KS
CP

Date/Time

Received by:

Relinquished by:

3/5/18 10:12 AM
CP
KS