



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93693

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/05/2018	10:25
2	J1792-12		1	>100.0	>212.0	03/05/2018	10:55
3	J1792-12DUP		1	>100.0	>212.0	03/05/2018	11:15
4	J1792-14		1	>100.0	>212.0	03/05/2018	11:35
5	J1792-16		1	>100.0	>212.0	03/05/2018	11:55

Result = (Celsius * 1.8) + 32

LB 93693

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP3518

WorkList ID : 109246

Date : 3/5/2018 8:10:30 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/09/2018	Solid	J1789-01	Flash Point	Cool 4 deg C	PSEG02	G31	19614	03/02/2018	1010A
03/07/2018	Water	J1792-12 C	Flash Point	Cool 4 deg C	PSEG01	G32	TP-7	03/01/2018	1010A
03/07/2018	Water	J1792-14 C	Flash Point	Cool 4 deg C	PSEG01	G32	TP-8	03/01/2018	1010A
03/07/2018	Water	J1792-16 C	Flash Point	Cool 4 deg C	PSEG01	G32	TP-13	03/01/2018	1010A

Date/Time 3/5/18 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 3/5/18 12:00 PM
Received by: CP
Relinquished by: KS