



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB93695

Temp.: 22.2 °C

SLOPE.: 99.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/05/2018	14:35
2	CAL2	1	Water			7.00	03/05/2018	14:37
3	CAL3	1	Water			10.01	03/05/2018	14:39
4	ICV	1	Water			7.00	03/05/2018	14:41
5	CCV1	1	Water			2.00	03/05/2018	14:43
6	J1804-01	1	Water			4.73	03/05/2018	14:45
7	J1804-01DUP	1	Water			4.74	03/05/2018	14:47
8	CCV2	1	Water			12.00	03/05/2018	14:50

WORKLIST(Hardcopy Internal Chain)

LB 93 695

WorkList Name : pH1804

WorkList ID : 109268

Date : 3/5/2018 2:21:56 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2018	Water	J1804-01	pH	Cool 4 deg C	STER04	H21	MAIN-CELL	03/02/2018	SM 4500-PH B

Date/Time 03/05/18 2:30 PM

Received by: SP

Relinquished by: CP

Date/Time 03/05/18 3:00 PM

Received by: CP

Relinquished by: SP