


7471B

LB 98700
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44671



A= 0.0000e+000

B= 2.1961e-004

C= -3.0632e-002

Rho= 0.9992377

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	54	0.000	54				
0.2	0.200	0.187	-0.013	990	0.0 %	990				
2.5	2.500	2.426	-0.074	11185	0.0 %	11185				
5.0	5.000	5.073	0.073	23239	0.0 %	23239				
7.5	7.500	7.754	0.254	35448	0.0 %	35448				
10	10.000	9.779	-0.221	44671	0.0 %	44671				

LB93700 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Mar 2018 15:25:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.00007471B	PPB		05 Mar 2018 15:29:48	S	Std
0.2 - 1	50.2	990	-	0.20007471B	PPB		05 Mar 2018 15:32:05	S	Std
2.5 - 1	52.5	11185	-	2.50007471B	PPB		05 Mar 2018 15:34:23	S	Std
5.0 - 1	55	23239	-	5.00007471B	PPB		05 Mar 2018 15:36:40	S	Std
7.5 - 1	57.5	35448	-	7.50007471B	PPB		05 Mar 2018 15:38:57	S	Std
10 - 1	510	44671	-	10.00007471B	PPB		05 Mar 2018 15:49:01	S	Std
ICV29 - 1	ICV29	19582	4.2697	-7471B	PPB		05 Mar 2018 15:54:28	U	SMPL
ICB29 - 1	ICB29	-208	-0.0763	-7471B	PPB		05 Mar 2018 15:56:44	U	SMPL
CCV91 - 1	CCV91	22721	4.9591	-7471B	PPB		05 Mar 2018 16:10:14	U	SMPL
CCB91 - 1	CCB91	-102	-0.0530	-7471B	PPB		05 Mar 2018 16:12:29	U	SMPL
CRA - 1	CRA	945	0.1769	-7471B	PPB		05 Mar 2018 16:14:46	U	SMPL
HighStd - 1	HighStd	45005	9.8528	-7471B	PPB		05 Mar 2018 16:22:00	U	SMPL
ChkStd - 1	ChkStd	31779	6.9483	-7471B	PPB		05 Mar 2018 16:24:16	U	SMPL
PB107047BL - 1	PB107047BL	-41	-0.0396	-7471B	PPB		05 Mar 2018 16:29:09	U	SMPL
PB107047BS - 1	PB107047BS	15045	3.2734	-7471B	PPB		05 Mar 2018 16:31:25	U	SMPL
J1761-01 - 1	TR-04-0302018	4475	0.9521	-7471B	PPB		05 Mar 2018 16:33:42	U	SMPL
J1783-01 - 1	SLUDGE	16404	3.5718	-7471B	PPB		05 Mar 2018 16:35:59	U	SMPL
J1786-01 - 1	SAMPLE-1-24+31-WEST	14609	3.1776	-7471B	PPB		05 Mar 2018 16:38:16	U	SMPL
J1786-03 - 1	SAMPLE-2-25+71-WEST	5390	1.1531	-7471B	PPB		05 Mar 2018 16:40:34	U	SMPL
J1792-01 - 1	TP-6	13821	3.0046	-7471B	PPB		05 Mar 2018 16:42:50	U	SMPL
CCV92 - 1	CCV92	21853	4.7685	-7471B	PPB		05 Mar 2018 16:45:06	U	SMPL
CCB92 - 1	CCB92	-149	-0.0634	-7471B	PPB		05 Mar 2018 16:47:22	U	SMPL
J1792-01DUP - 1	TP-6DUP	12489	2.7120	-7471B	PPB		05 Mar 2018 16:49:38	U	SMPL
J1792-01MS - 1	TP-6MS	30781	6.7291	-7471B	PPB		05 Mar 2018 16:51:54	U	SMPL
J1792-01MSD - 1	TP-6MSD	29395	6.4247	-7471B	PPB		05 Mar 2018 16:54:10	U	SMPL
J1792-03 - 1	TP-7	21500	4.6909	-7471B	PPB		05 Mar 2018 16:56:27	U	SMPL
J1792-05 - 1	TP-8	18854	4.1099	-7471B	PPB		05 Mar 2018 16:58:45	U	SMPL
J1792-07 - 1	TP-12	2005	0.4097	-7471B	PPB		05 Mar 2018 17:01:02	U	SMPL
J1792-09 - 1	TP-13	12924	2.8076	-7471B	PPB		05 Mar 2018 17:03:20	U	SMPL
J1797-01 - 1	SI-1	2054	0.4204	-7471B	PPB		05 Mar 2018 17:05:38	U	SMPL
J1797-02 - 1	SI-2	909	0.1690	-7471B	PPB		05 Mar 2018 17:07:55	U	SMPL
J1797-03 - 1	SI-3	1487	0.2959	-7471B	PPB		05 Mar 2018 17:10:11	U	SMPL
CCV93 - 1	CCV93	22274	4.8609	-7471B	PPB		05 Mar 2018 17:12:27	U	SMPL
CCB93 - 1	CCB93	-141	-0.0616	-7471B	PPB		05 Mar 2018 17:14:43	U	SMPL
J1798-01 - 1	TRK-STONE	235	0.0210	-7471B	PPB		05 Mar 2018 17:17:00	U	SMPL
J1792-01LX5 - 1		2534	0.5259	-7471B	PPB		05 Mar 2018 17:19:16	U	SMPL
J1792-01A - 1		30216	6.6050	-7471B	PPB		05 Mar 2018 17:21:32	U	SMPL
CCV94 - 1	CCV94	21986	4.7977	-7471B	PPB		05 Mar 2018 17:23:49	U	SMPL
CCB94 - 1	CCB94	-94	-0.0513	-7471B	PPB		05 Mar 2018 17:26:06	U	SMPL