



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93705

Temp.: 21.2 °C

BalanceID: WC SC-4

SLOPE.: 97.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/05/2018	17:50
2	CAL2	1	Water			7.01	03/05/2018	17:51
3	CAL3	1	Water			10.02	03/05/2018	18:00
4	ICV	1	Water			7.00	03/05/2018	18:01
5	CCV1	1	Water			2.02	03/05/2018	18:02
6	J1731-06	1	Solid	20.03	20	9.02	03/05/2018	18:03
7	J1800-06	1	Solid	20.01	20	6.17	03/05/2018	18:05
8	J1802-01	1	Solid	20.02	20	5.78	03/05/2018	18:06
9	J1808-03	1	Solid	20.04	20	8.90	03/05/2018	18:08
10	J1809-04	1	Solid	20.03	20	8.75	03/05/2018	18:10
11	J1809-08	1	Solid	20.05	20	8.67	03/05/2018	18:11
12	J1810-02	1	Solid	20.03	20	7.61	03/05/2018	18:12
13	J1810-04	1	Solid	20.02	20	7.03	03/05/2018	18:15
14	J1810-04DUP	1	Solid	20.04	20	7.04	03/05/2018	18:16
15	CCV2	1	Water			12.02	03/05/2018	18:19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J1809

WorkList ID : 109274

Date : 3/5/2018 5:40:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2018	Solid	J1731-06	Corrosivity	Cool 4 deg C	PSEG02		BALER-SOIL	03/05/2018	9045C
03/05/2018	Solid	J1800-06	Corrosivity	Cool 4 deg C	PSEG01	I11	OILY-DEBRIS	03/05/2018	9045C
03/05/2018	Solid	J1802-01	Corrosivity	Cool 4 deg C	PSEG02	I12	DRUM-1-2	03/05/2018	9045C
03/05/2018	Solid	J1808-03	Corrosivity	Cool 4 deg C	GZAG04	H32	WC-02-01-17	03/05/2018	9045C
03/05/2018	Solid	J1809-04	Corrosivity	Cool 4 deg C	PSEG01	I12	RO-1-4	03/05/2018	9045C
03/05/2018	Solid	J1809-08	Corrosivity	Cool 4 deg C	PSEG01	I12	RO-5-7	03/05/2018	9045C
03/05/2018	Solid	J1810-02	Corrosivity	Cool 4 deg C	PSEG01	I13	MHA-WCS	03/05/2018	9045C
03/05/2018	Solid	J1810-04	Corrosivity	Cool 4 deg C	PSEG01	I13	MHA-WCD	03/05/2018	9045C

Date/Time

Received by:

Relinquished by:

03-05-18 5:30 PM

SP

SP

Date/Time

Received by:

Relinquished by:

03-05-18 6:35 PM

CD

JB