



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93706

Temp.: 16.5 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/05/2018	17:50
2	CAL2	1	Water			7.01	03/05/2018	17:51
3	CAL3	1	Water			10.03	03/05/2018	18:00
4	ICV	1	Water			7.00	03/05/2018	18:01
5	CCV1	1	Water			2.02	03/05/2018	18:02
6	J1810-06	1	Water			6.90	03/05/2018	18:03
7	J1810-06DUP	1	Water			6.91	03/05/2018	18:05
8	CCV2	1	Water			12.01	03/05/2018	18:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1810 WorkList ID : 109275 Date : 3/5/2018 5:40:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2018	Water	J1810-06	-F Corrosivity	Cool 4 deg C	PSEG01	I13	MHA-WCGW	03/05/2018	9040C

Date/Time 03-05-18 5:41 PM
 Received by: JS
 Relinquished by: JS

Date/Time 03-05-18 6:40 PM
 Received by: JS
 Relinquished by: JS