



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: Hiral

Run Number: LB93713

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/06/2018	09:10
2	J1810-06		1	>100.0	>212.0	03/06/2018	09:40
3	J1810-06DUP		1	>100.0	>212.0	03/06/2018	10:05

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP3618

WorkList ID : 109290

Date : 3/6/2018 8:18:10 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/08/2018	Water	J1810-06	C Flash Point	Cool 4 deg C	PSEG01	I13	MHA-WCGW	03/05/2018	1010A

Date/Time

Received by:

Relinquished by:

3/6/18 8:30 AM
KS
CP

Date/Time

Received by:

Relinquished by:

3/6/18 10:10 AM
CP
KS