


ISM02.4

LB 93 718
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42697



A= 0.0000e+000

B= 2.3437e-004

C= -1.4269e-002

Rho= 0.9999041

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.005	-0.005	39	0.000	39					-
0.2	0.200	0.227	0.027	1028	0.0 %	1028					14
2.5	2.500	2.518	0.018	10805	0.0 %	10805					1
5.0	5.000	4.900	-0.100	20969	0.0 %	20969					-2
7.5	7.500	7.568	0.068	32351	0.0 %	32351					1
10.0	10.000	9.992	-0.008	42697	0.0 %	42697					0

no3
3/6/18

LB93718
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	39	-		0 ISM02.4	PPB	3/6/2018 13:53	Std
	0.2 S0.2	1028	-		0.2 ISM02.4	PPB	3/6/2018 13:56	Std
	2.5 S2.5	10805	-		2.5 ISM02.4	PPB	3/6/2018 13:58	Std
	5 S5	20969	-		5 ISM02.4	PPB	3/6/2018 14:00	Std
	7.5 S7.5	32351	-		7.5 ISM02.4	PPB	3/6/2018 14:03	Std
	10 S10	42697	-		10 ISM02.4	PPB	3/6/2018 14:05	Std
ICV	ICV	16388	3.8265 -		ISM02.4	PPB	3/6/2018 14:08	SMPL
ICB	ICB	-142	-0.0475 -		ISM02.4	PPB	3/6/2018 14:10	SMPL
CCV002	CCV002	21264	4.9693 -		ISM02.4	PPB	3/6/2018 14:12	SMPL
CCB002	CCB002	-159	-0.0515 -		ISM02.4	PPB	3/6/2018 14:15	SMPL
PB107101BL	PBW003	-125	-0.0436 -		ISM02.4	PPB	3/6/2018 14:17	SMPL
J1729-01	MC0AC8	-116	-0.0415 -		ISM02.4	PPB	3/6/2018 14:19	SMPL
J1729-02	MC0AD0	-152	-0.0499 -		ISM02.4	PPB	3/6/2018 14:21	SMPL
J1729-03	MC0AD5	-18	-0.0185 -		ISM02.4	PPB	3/6/2018 14:24	SMPL
J1729-04	MC0AD5D	-58	-0.0279 -		ISM02.4	PPB	3/6/2018 14:26	SMPL
J1729-05	MC0AD5S	4373	1.0106 -		ISM02.4	PPB	3/6/2018 14:28	SMPL
J1729-06	MC0AD8	217	0.0366 -		ISM02.4	PPB	3/6/2018 14:31	SMPL
J1729-07	MC0AD9	-112	-0.0405 -		ISM02.4	PPB	3/6/2018 14:33	SMPL
J1729-08	MC0AE0	-109	-0.0398 -		ISM02.4	PPB	3/6/2018 14:35	SMPL
J1729-09	MC0AE1	-63	-0.029 -		ISM02.4	PPB	3/6/2018 14:37	SMPL
J1729-10	MC0AE2	-91	-0.0356 -		ISM02.4	PPB	3/6/2018 14:40	SMPL
J1729-11	MC0AE3	-45	-0.0248 -		ISM02.4	PPB	3/6/2018 14:42	SMPL
J1729-12	MC0AE4	-111	-0.0403 -		ISM02.4	PPB	3/6/2018 14:44	SMPL
J1729-13	MC0AC9	-103	-0.0384 -		ISM02.4	PPB	3/6/2018 14:47	SMPL
J1729-14	MC0AD1	-103	-0.0384 -		ISM02.4	PPB	3/6/2018 14:49	SMPL
J1729-15	MC0AD2	-82	-0.0335 -		ISM02.4	PPB	3/6/2018 14:51	SMPL
J1729-16	MC0AD3	-105	-0.0389 -		ISM02.4	PPB	3/6/2018 14:53	SMPL
J1729-17	MC0AD4	-79	-0.0328 -		ISM02.4	PPB	3/6/2018 14:56	SMPL
J1729-18	MC0AE5	-92	-0.0358 -		ISM02.4	PPB	3/6/2018 14:58	SMPL
PB107102BL	PBW003	-20	-0.019 -		ISM02.4	PPB	3/6/2018 15:00	SMPL
CCV003	CCV003	20611	4.8163 -		ISM02.4	PPB	3/6/2018 15:03	SMPL
CCB003	CCB003	-133	-0.0454 -		ISM02.4	PPB	3/6/2018 15:05	SMPL
J1743-01	MC0B91	-74	-0.0316 -		ISM02.4	PPB	3/6/2018 15:07	SMPL
J1743-02	MC0B92	-84	-0.034 -		ISM02.4	PPB	3/6/2018 15:09	SMPL
J1743-03	MC0B93	-71	-0.0309 -		ISM02.4	PPB	3/6/2018 15:12	SMPL
J1743-04	MC0B93D	-69	-0.0304 -		ISM02.4	PPB	3/6/2018 15:14	SMPL
J1743-05	MC0B93S	4191	0.968 -		ISM02.4	PPB	3/6/2018 15:16	SMPL
J1743-06	MC0B94	-118	-0.0419 -		ISM02.4	PPB	3/6/2018 15:18	SMPL
J1743-07	MC0B95	-66	-0.0297 -		ISM02.4	PPB	3/6/2018 15:21	SMPL
J1743-08	MC0B96	-73	-0.0314 -		ISM02.4	PPB	3/6/2018 15:23	SMPL
J1743-09	MC0B97	-103	-0.0384 -		ISM02.4	PPB	3/6/2018 15:25	SMPL
J1743-10	MC0B98	-63	-0.029 -		ISM02.4	PPB	3/6/2018 15:28	SMPL
J1743-11	MC0B99	-65	-0.0295 -		ISM02.4	PPB	3/6/2018 15:30	SMPL
J1743-12	MC0BA0	-70	-0.0307 -		ISM02.4	PPB	3/6/2018 15:32	SMPL
J1743-13	MC0AE6	-79	-0.0328 -		ISM02.4	PPB	3/6/2018 15:34	SMPL
J1743-14	MC0AE7	-104	-0.0386 -		ISM02.4	PPB	3/6/2018 15:37	SMPL

LB93718

INSTRUMENT ID : CV1

J1743-15	MCOAE8	-64	-0.0293 -	ISM02.4	PPB	3/6/2018 15:39	SMPL
J1743-16	MCOAE9	-77	-0.0323 -	ISM02.4	PPB	3/6/2018 15:41	SMPL
J1743-17	MCOAF0	-88	-0.0349 -	ISM02.4	PPB	3/6/2018 15:43	SMPL
J1743-18	MCOAF1	-97	-0.037 -	ISM02.4	PPB	3/6/2018 15:46	SMPL
CCV004	CCV004	20786	4.8573 -	ISM02.4	PPB	3/6/2018 15:48	SMPL
CCB004	CCB004	-149	-0.0492 -	ISM02.4	PPB	3/6/2018 15:50	SMPL