


245.1

LB93724
INSTRUMENT ID: CY2

Linear

 μ Abs.:

33194



A= 0.0000e+000

B= 3.0314e-004

C= -5.5789e-002

Rho= 0.9998573

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.040	-0.040	52	0.000	52				
0.05	0.050	0.058	0.008	376	0.0 %	376				
0.2	0.200	0.180	-0.020	777	0.0 %	777				
2.5	2.500	2.526	0.026	8517	0.0 %	8517				
5.0	5.000	5.120	0.120	17075	0.0 %	17075				
7.5	7.500	7.399	-0.101	24591	0.0 %	24591				
10.0	10.000	10.007	0.007	33194	0.0 %	33194				

LB93724 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 06 Mar 2018 16:18:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.0000	245.1	PPB	06 Mar 2018 16:20:57	S	Std
0.05 - 1	50.05	376	-	0.0500	245.1	PPB	06 Mar 2018 16:23:06	S	Std
0.2 - 1	50.2	777	-	0.2000	245.1	PPB	06 Mar 2018 16:25:15	S	Std
2.5 - 1	52.5	8517	-	2.5000	245.1	PPB	06 Mar 2018 16:27:23	S	Std
5.0 - 1	55	17075	-	5.0000	245.1	PPB	06 Mar 2018 16:29:31	S	Std
7.5 - 1	57.5	24591	-	7.5000	245.1	PPB	06 Mar 2018 16:31:38	S	Std
10.0 - 1	510	33194	-	10.0000	245.1	PPB	06 Mar 2018 16:33:46	S	Std
ICV31 - 1	ICV31	13041	3.8975	-245.1		PPB	06 Mar 2018 16:36:37	U	SMPL
ICB31 - 1	ICB31	-74	-0.0782	-245.1		PPB	06 Mar 2018 16:38:43	U	SMPL
CCV05 - 1	CCV05	16587	4.9725	-245.1		PPB	06 Mar 2018 16:40:51	U	SMPL
CCB05 - 1	CCB05	-93	-0.0840	-245.1		PPB	06 Mar 2018 16:42:57	U	SMPL
CRA - 1	CRA	356	0.0521	-245.1		PPB	06 Mar 2018 16:45:05	U	SMPL
HighStd - 1	HighStd	33245	10.0222	-245.1		PPB	06 Mar 2018 16:47:12	U	SMPL
ChkStd - 1	ChkStd	23279	7.0011	-245.1		PPB	06 Mar 2018 16:49:19	U	SMPL
PB107100BL - 1	PBW	-74	-0.0782	-245.1		PPB	06 Mar 2018 16:56:17	U	SMPL
PB107100BS - 1	LCSW	12568	3.7541	-245.1		PPB	06 Mar 2018 16:58:23	U	SMPL
J1794-01 - 1	030118-SIDI-F-D-M	-162	-0.1049	-245.1		PPB	06 Mar 2018 17:02:57	U	SMPL
J1794-02 - 1	030118-SIDI-F-U-M	-176	-0.1091	-245.1		PPB	06 Mar 2018 17:05:05	U	SMPL
J1794-03 - 1	030118-JETT-F-D-M	-172	-0.1079	-245.1		PPB	06 Mar 2018 17:07:13	U	SMPL
J1794-04 - 1	030118-JETT-F-U-M	-177	-0.1094	-245.1		PPB	06 Mar 2018 17:09:22	U	SMPL
J1796-01 - 1	COMP	-84	-0.0813	-245.1		PPB	06 Mar 2018 17:11:28	U	SMPL
CCV06 - 1	CCV06	16858	5.0546	-245.1		PPB	06 Mar 2018 17:13:35	U	SMPL
CCB06 - 1	CCB06	-74	-0.0782	-245.1		PPB	06 Mar 2018 17:15:42	U	SMPL
J1796-01DUP - 1	COMPDUP	-86	-0.0819	-245.1		PPB	06 Mar 2018 17:17:50	U	SMPL
J1796-01MS - 1	COMPMS	12631	3.7732	-245.1		PPB	06 Mar 2018 17:19:57	U	SMPL
J1796-01MSD - 1	COMPMSD	13324	3.9833	-245.1		PPB	06 Mar 2018 17:22:04	U	SMPL
J1796-01LX5 - 1		-66	-0.0758	-245.1		PPB	06 Mar 2018 17:24:12	U	SMPL
J1796-01A - 1		13374	3.9985	-245.1		PPB	06 Mar 2018 17:26:19	U	SMPL
CCV07 - 1	CCV07	16771	5.0282	-245.1		PPB	06 Mar 2018 17:28:27	U	SMPL
CCB07 - 1	CCB07	-62	-0.0746	-245.1		PPB	06 Mar 2018 17:30:35	U	SMPL