



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 03/08/2018DATE & TIME IN: 03/06/2018 16:30OVEN TEMP: 103 °CRun Number: LB93725DATE & TIME OUT: 03/08/2018 08:00OVEN TEMP: 103 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB93725BL	LB93725BL	82.4808	82.4808	100	82.4808	82.4808	82.4808	0	0
2	J1814-01	DREDGE-DECANT-WATER	74.8284	74.8284	100	75.0964	75.0964	75.0964	0.268	2680
3	J1814-01DUP	DREDGE-DECANT-WATERDUP	90.2887	90.2887	100	90.5614	90.5614	90.5614	0.2727	2727

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS030618

WorkList ID : 109327

Date : 3/6/2018 4:16:58 PM

LB93725

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/09/2018	Water	J1814-01	TS	Cool 4 deg C	TULL01	H33	DREDGE-DECANT-WATER	03/05/2018	SM2540 B

Date/Time 03/06/18, 16:00
 Received by: AS
 Relinquished by: CB

Date/Time 03/06/18, 16:45
 Received by: CB
 Relinquished by: AS