



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93727

Temp.: 15.3 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	03/06/2018	16:00
2	CAL2	1	Water			7.01	03/06/2018	16:01
3	CAL3	1	Water			10.02	03/06/2018	16:02
4	ICV	1	Water			6.99	03/06/2018	16:04
5	CCV1	1	Water			2.02	03/06/2018	16:05
6	J1819-08	1	Water			7.13	03/06/2018	16:06
7	J1819-08DUP	1	Water			7.14	03/06/2018	16:08
8	CCV2	1	Water			12.02	03/06/2018	16:10

2B93f2f

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J1819 WorkList ID : 109325 Date : 3/6/2018 3:34:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/06/2018	Water	J1819-08	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-10	03/06/2018	9040C

Date/Time 03-06-18 4:00 PM
Received by: Jg
Relinquished by: G

Date/Time 03-06-18 6:00 PM
Received by: Jg
Relinquished by: Jg