


7471B

LB93736
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43652



A= 0.0000e+000

B= 2.2958e-004

C= -2.4491e-003

Rho= 0.9997917

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	170	0.000	170				
0.2	0.200	0.237	0.037	1043	0.0 %	1043				
2.5	2.500	2.486	-0.014	10839	0.0 %	10839				
5.0	5.000	4.843	-0.157	21106	0.0 %	21106				
7.5	7.500	7.578	0.078	33020	0.0 %	33020				
10	10.000	10.019	0.019	43652	0.0 %	43652				

LB93736 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 06 Mar 2018 19:26:12				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	170	-	0.0000	7471B	PPB	06 Mar 2018 19:28:05	S Std
0.2 - 1	50.2	1043	-	0.2000	7471B	PPB	06 Mar 2018 19:30:22	S Std
2.5 - 1	52.5	10839	-	2.5000	7471B	PPB	06 Mar 2018 19:32:40	S Std
5.0 - 1	55	21106	-	5.0000	7471B	PPB	06 Mar 2018 19:34:57	S Std
7.5 - 1	57.5	33020	-	7.5000	7471B	PPB	06 Mar 2018 19:37:15	S Std
10 - 1	510	43652	-	10.0000	7471B	PPB	06 Mar 2018 19:39:32	S Std
ICV32 - 1	ICV32	15693	3.6003	-	7471B	PPB	06 Mar 2018 19:43:48	U SMPL
ICB32 - 1	ICB32	21	0.0024	-	7471B	PPB	06 Mar 2018 19:46:04	U SMPL
CCV08 - 1	CCV08	21275	4.8819	-	7471B	PPB	06 Mar 2018 19:48:19	U SMPL
CCB08 - 1	CCB08	-24	-0.0080	-	7471B	PPB	06 Mar 2018 19:50:35	U SMPL
CRA - 1	CRA	1060	0.2409	-	7471B	PPB	06 Mar 2018 19:52:52	U SMPL
HighStd - 1	HighStd	43428	9.9677	-	7471B	PPB	06 Mar 2018 19:55:08	U SMPL
ChkStd - 1	ChkStd	30054	6.8973	-	7471B	PPB	06 Mar 2018 19:57:24	U SMPL
PB107106BL - 1	PBS	-87	-0.0224	-	7471B	PPB	06 Mar 2018 19:59:41	U SMPL
PB107106BS - 1	LCSS	14945	3.4286	-	7471B	PPB	06 Mar 2018 20:02:00	U SMPL
J1731-04 - 1	BALER-SOIL	7066	1.6198	-	7471B	PPB	06 Mar 2018 20:04:16	U SMPL
J1808-02 - 1	SB-04-9.5-11	15625	3.5847	-	7471B	PPB	06 Mar 2018 20:06:34	U SMPL
J1808-05 - 1	SB-06-05-10	45648	10.4774	-	7471B	PPB	06 Mar 2018 20:08:52	U SMPL
J1809-01 - 1	RO-1-4	7626	1.7483	-	7471B	PPB	06 Mar 2018 20:11:09	U SMPL
J1809-01DUP - 1	RO-1-4DUP	8900	2.0408	-	7471B	PPB	06 Mar 2018 20:13:27	U SMPL
CCV09 - 1	CCV09	20997	4.8180	-	7471B	PPB	06 Mar 2018 20:15:43	U SMPL
CCB09 - 1	CCB09	11	0.0001	-	7471B	PPB	06 Mar 2018 20:18:00	U SMPL
J1809-01MS - 1	RO-1-4MS	27568	6.3266	-	7471B	PPB	06 Mar 2018 20:20:15	U SMPL
J1809-01MSD - 1	RO-1-4MSD	23515	5.3961	-	7471B	PPB	06 Mar 2018 20:25:13	U SMPL
J1809-05 - 1	RO-5-7	1163	0.2646	-	7471B	PPB	06 Mar 2018 20:27:29	U SMPL
J1809-09 - 1	CONCRETE	647	0.1461	-	7471B	PPB	06 Mar 2018 20:29:46	U SMPL
J1810-01 - 1	MHA-WCS	62176	14.2719	-	7471B	PPB	06 Mar 2018 20:32:03	U SMPL
J1810-03 - 1	MHA-WCD	3188	0.7294	-	7471B	PPB	06 Mar 2018 20:34:20	U SMPL
J1822-01 - 1	ELIZ-OILY-DERBIS	459	0.1029	-	7471B	PPB	06 Mar 2018 20:36:40	U SMPL
J1809-01LX5 - 1		1556	0.3548	-	7471B	PPB	06 Mar 2018 20:38:57	U SMPL
J1809-01A - 1		25022	5.7421	-	7471B	PPB	06 Mar 2018 20:41:15	U SMPL
J1808-05DLX2 - 1		22523	5.1684	-	7471B	PPB	06 Mar 2018 20:43:31	U SMPL
CCV10 - 1	CCV10	21087	4.8387	-	7471B	PPB	06 Mar 2018 20:45:47	U SMPL
CCB10 - 1	CCB10	21	0.0024	-	7471B	PPB	06 Mar 2018 20:48:04	U SMPL
J1810-01DLX2 - 1		30686	7.0424	-	7471B	PPB	06 Mar 2018 20:50:21	U SMPL
CCV11 - 1	CCV11	21242	4.8743	-	7471B	PPB	06 Mar 2018 20:52:37	U SMPL
CCB11 - 1	CCB11	-1	-0.0027	-	7471B	PPB	06 Mar 2018 20:54:55	U SMPL