



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93739

Temp.: 14.6 °C

SLOPE.: 97.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/07/2018	12:25
2	CAL2	1	Water			7.01	03/07/2018	12:26
3	CAL3	1	Water			10.02	03/07/2018	12:27
4	ICV	1	Water			7.00	03/07/2018	12:30
5	CCV1	1	Water			2.01	03/07/2018	12:31
6	J1833-01	1	Water			9.16	03/07/2018	12:32
7	J1833-01DUP	1	Water			9.17	03/07/2018	12:33
8	CCV2	1	Water			12.02	03/07/2018	12:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1833

WorkList ID : 109333

Date : 3/7/2018 12:20:56 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/06/2018	Water	J1833-01-E	pH	Cool 4 deg C	VIAS01	H11	EFF-180306-02	03/06/2018	SM 4500-PH B

Date/Time 03-07-18 12:20 PM
 Received by: Jo
 Relinquished by: CP

Date/Time 03-07-18 11:00 AM
 Received by: CP
 Relinquished by: Jo