

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/08/2018

DATE & TIME IN: 03/08/2018 10:00

OVEN TEMP: 103 °C

Run Number: LB93743

DATE & TIME OUT: 03/08/2018 13:45

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB93743BL	LB93743BL	1.8111	1.8111	100	1.8111	1.8111	1.8111	0	0
2	LB93743BS	LB93743BS	1.8697	1.8697	100	1.9258	1.9258	1.9258	0.0561	561
3	J1817-01	030518-JETT-F-D-B	1.8971	1.8971	200	1.9125	1.9125	1.9125	0.0154	77
4	J1817-02	030518-JETT-F-D-S	1.8937	1.8937	200	1.9047	1.9047	1.9047	0.011	55
5	J1817-03	030518-JETT-F-U-B	1.9099	1.9099	200	1.9228	1.9228	1.9228	0.0129	64.5
6	J1817-04	030518-JETT-F-U-S	1.9274	1.9274	200	1.987	1.987	1.987	0.0596	298
7	J1819-07	TP-10	1.9041	1.9041	100	1.9282	1.9282	1.9282	0.0241	241
8	J1819-07DUP	TP-10DUP	1.8942	1.8942	100	1.919	1.919	1.919	0.0248	248
9	J1827-01	COMPOSITE-A-TOTAL	1.8911	1.8911	300	1.9153	1.9153	1.9153	0.0242	80.7
10	J1827-02	COMPOSITE-B-TOTAL	1.8878	1.8878	300	1.9085	1.9085	1.9085	0.0207	69
11	J1833-01	EFF-180306-02	1.8325	1.8325	800	1.8332	1.8332	1.8332	0.0007	0.9
12	J1834-02	OUTFALL001-180306	1.821	1.821	800	1.8215	1.8215	1.8215	0.0005	0.6

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS030818

WorkList ID : 109354

Date : 3/8/2018 8:25:04 AM

1193743

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2018	Water	J1817-01 C	TSS	Cool 4 deg C	HDR103	H43	030518-JETT-F-D-B	03/05/2018	SM2540 D
03/12/2018	Water	J1817-02 C	TSS	Cool 4 deg C	HDR103	H43	030518-JETT-F-D-S	03/05/2018	SM2540 D
03/12/2018	Water	J1817-03 C	TSS	Cool 4 deg C	HDR103	H43	030518-JETT-F-U-B	03/05/2018	SM2540 D
03/12/2018	Water	J1817-04 C	TSS	Cool 4 deg C	HDR103	H43	030518-JETT-F-U-S	03/05/2018	SM2540 D
03/09/2018	Water	J1819-07 G	TSS	Cool 4 deg C	PSEG01	I13	TP-10	03/05/2018	SM2540 D
02/21/2018	Water	J1827-01 C	TSS	Cool 4 deg C	AQUA04	H12	COMPOSITE-A-TOTAL	02/14/2018	SM2540 D
02/21/2018	Water	J1827-02 C	TSS	Cool 4 deg C	AQUA04	H12	COMPOSITE-B-TOTAL	02/14/2018	SM2540 D
03/09/2018	Water	J1833-01 D	TSS	Cool 4 deg C	VIAS01	H11	EFF-180306-02	03/06/2018	SM2540 D
03/13/2018	Water	J1834-02 E	TSS	Cool 4 deg C	SHAW22	H11	OUTFALL001-180306	03/06/2018	SM2540 D

Date/Time 03/08/18, 09:00
 Received by: AS
 Relinquished by: CS

Date/Time 03/08/18, 10:30
 Received by: CS
 Relinquished by: AS