


7470A

LB93749
INSTRUMENT ID: LV1

Linear

 μ Abs.:

36469



A= 0.0000e+000

B= 2.7430e-004

C= 1.3154e-002

Rho= 0.9999880

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	-65	0.000	-65				
0.2	0.200	0.225	0.025	774	0.0 %	774				
2.5	2.500	2.494	-0.006	9045	0.0 %	9045				
5.0	5.000	4.969	-0.031	18066	0.0 %	18066				
7.5	7.500	7.499	-0.001	27292	0.0 %	27292				
10.0	10.000	10.017	0.017	36469	0.0 %	36469				

LB93749 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Mar 2018 12:09:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-65	-	0.00007470A	PPB		08 Mar 2018 12:12:54	S	Std
0.2 - 1	50.2	774	-	0.20007470A	PPB		08 Mar 2018 12:15:02	S	Std
2.5 - 1	52.5	9045	-	2.50007470A	PPB		08 Mar 2018 12:17:10	S	Std
5.0 - 1	55	18066	-	5.00007470A	PPB		08 Mar 2018 12:19:19	S	Std
7.5 - 1	57.5	27292	-	7.50007470A	PPB		08 Mar 2018 12:21:27	S	Std
10.0 - 1	510	36469	-	10.00007470A	PPB		08 Mar 2018 12:23:35	S	Std
ICV33 - 1	ICV33	13724	3.7777	-7470A	PPB		08 Mar 2018 12:26:17	U	SMPL
ICB33 - 1	ICB33	-188	-0.0384	-7470A	PPB		08 Mar 2018 12:28:23	U	SMPL
CCV12 - 1	CCV12	17914	4.9270	-7470A	PPB		08 Mar 2018 12:30:30	U	SMPL
CCB12 - 1	CCB12	-245	-0.0541	-7470A	PPB		08 Mar 2018 12:32:37	U	SMPL
CRA - 1	CRA	710	0.2079	-7470A	PPB		08 Mar 2018 12:34:45	U	SMPL
HighStd - 1	HighStd	36804	10.1087	-7470A	PPB		08 Mar 2018 12:36:51	U	SMPL
ChkStd - 1	ChkStd	24567	6.7520	-7470A	PPB		08 Mar 2018 12:38:58	U	SMPL
PB107133BL - 1	PBW	-346	-0.0818	-7470A	PPB		08 Mar 2018 12:41:08	U	SMPL
PB107133BS - 1	LCSW	13895	3.8246	-7470A	PPB		08 Mar 2018 12:43:17	U	SMPL
PB107133TB - 1	PB107103TB	-165	-0.0321	-7470A	PPB		08 Mar 2018 12:45:24	U	SMPL
J1731-06 - 1	BALER-SOIL	-71	-0.0063	-7470A	PPB		08 Mar 2018 12:47:33	U	SMPL
J1731-06DUP - 1	BALER-SOILDUP	-64	-0.0044	-7470A	PPB		08 Mar 2018 12:49:41	U	SMPL
J1731-06MS - 1	BALER-SOILMS	12928	3.5594	-7470A	PPB		08 Mar 2018 12:51:50	U	SMPL
J1731-06MSD - 1	BALER-SOILMSD	12560	3.4584	-7470A	PPB		08 Mar 2018 12:53:57	U	SMPL
CCV13 - 1	CCV13	18180	5.0000	-7470A	PPB		08 Mar 2018 12:56:03	U	SMPL
CCB13 - 1	CCB13	-254	-0.0565	-7470A	PPB		08 Mar 2018 12:58:10	U	SMPL
J1800-02 - 1	OIL	-138	-0.0247	-7470A	PPB		08 Mar 2018 13:00:18	U	SMPL
J1800-06 - 1	OILY-DEBRIS	-30	0.0049	-7470A	PPB		08 Mar 2018 13:02:25	U	SMPL
J1802-01 - 1	DRUM-1-2	-37	0.0030	-7470A	PPB		08 Mar 2018 13:04:31	U	SMPL
J1808-03 - 1	WC-02-01-17	-16	0.0088	-7470A	PPB		08 Mar 2018 13:06:38	U	SMPL
J1809-04 - 1	RO-1-4	20	0.0186	-7470A	PPB		08 Mar 2018 13:08:46	U	SMPL
J1809-08 - 1	RO-5-7	-56	-0.0022	-7470A	PPB		08 Mar 2018 13:10:53	U	SMPL
J1809-10 - 1	CONCRETE	-65	-0.0047	-7470A	PPB		08 Mar 2018 13:13:02	U	SMPL
J1810-02 - 1	MHA-WCS	-30	0.0049	-7470A	PPB		08 Mar 2018 13:15:10	U	SMPL
J1810-04 - 1	MHA-WCD	174	0.0609	-7470A	PPB		08 Mar 2018 13:17:18	U	SMPL
PB107134BL - 1	PBW	-154	-0.0291	-7470A	PPB		08 Mar 2018 13:19:25	U	SMPL
CCV14 - 1	CCV14	17742	4.8799	-7470A	PPB		08 Mar 2018 13:21:33	U	SMPL
CCB14 - 1	CCB14	-175	-0.0348	-7470A	PPB		08 Mar 2018 13:23:40	U	SMPL
PB107134BS - 1	LCSW	13660	3.7602	-7470A	PPB		08 Mar 2018 13:25:48	U	SMPL
PB107134TB - 1	PB107105TB	-185	-0.0376	-7470A	PPB		08 Mar 2018 13:27:54	U	SMPL
J1810-06 - 1	MHA-WCGW	-75	-0.0074	-7470A	PPB		08 Mar 2018 13:30:02	U	SMPL
J1810-06DUP - 1	MHA-WCGWDUP	-58	-0.0028	-7470A	PPB		08 Mar 2018 13:32:09	U	SMPL
J1810-06MS - 1	MHA-WCGWMS	13204	3.6351	-7470A	PPB		08 Mar 2018 13:34:17	U	SMPL
J1810-06MSD - 1	MHA-WCGWMSD	13009	3.5816	-7470A	PPB		08 Mar 2018 13:36:24	U	SMPL
PB107135BL - 1	PBW	-242	-0.0532	-7470A	PPB		08 Mar 2018 13:38:33	U	SMPL
PB107135BS - 1	LCSW	14157	3.8965	-7470A	PPB		08 Mar 2018 13:40:42	U	SMPL
J1810-05DLX10 - 1	MHA-WCGW	110072	30.2064	-7470A	PPB		08 Mar 2018 13:56:20	U	SMPL
J1810-05DUPDLX10 - 1	MHA-WCGWDUP	112204	30.7912	-7470A	PPB		08 Mar 2018 13:58:26	U	SMPL
CCV15 - 1	CCV15	17793	4.8939	-7470A	PPB		08 Mar 2018 14:00:39	U	SMPL
CCB15 - 1	CCB15	-334	-0.0785	-7470A	PPB		08 Mar 2018 14:02:53	U	SMPL
J1810-05MSDLX10 - 1	MHA-WCGWMS	112162	30.7797	-7470A	PPB		08 Mar 2018 14:05:02	U	SMPL
J1810-05MSDDLX10 - 1	MHA-WCGWMSD	113464	31.1368	-7470A	PPB		08 Mar 2018 14:07:09	U	SMPL
J1731-06LX5 - 1		-548	-0.1372	-7470A	PPB		08 Mar 2018 14:09:22	U	SMPL
J1731-06A - 1		13431	3.6973	-7470A	PPB		08 Mar 2018 14:11:33	U	SMPL
J1810-06LX5 - 1		-190	-0.0390	-7470A	PPB		08 Mar 2018 14:13:41	U	SMPL
J1810-06A - 1		13458	3.7047	-7470A	PPB		08 Mar 2018 14:15:49	U	SMPL
J1810-05LX50 - 1		22805	6.2687	-7470A	PPB		08 Mar 2018 14:17:57	U	SMPL
J1810-05ADLX10 - 1		115924	31.8116	-7470A	PPB		08 Mar 2018 14:20:06	U	SMPL
CCV16 - 1	CCV16	18094	4.9764	-7470A	PPB		08 Mar 2018 14:22:15	U	SMPL
CCB16 - 1	CCB16	-248	-0.0549	-7470A	PPB		08 Mar 2018 14:24:27	U	SMPL
J1810-05DLX50 - 1		20641	5.6751	-7470A	PPB		08 Mar 2018 14:26:35	U	SMPL
J1810-05DUPDLX50 - 1		21178	5.8224	-7470A	PPB		08 Mar 2018 14:28:42	U	SMPL
J1810-05MSDLX50 - 1		21239	5.8391	-7470A	PPB		08 Mar 2018 14:30:50	U	SMPL
J1810-05MSDDLX50 - 1		21004	5.7746	-7470A	PPB		08 Mar 2018 14:32:58	U	SMPL
J1810-05LDLX250 - 1		4039	1.1211	-7470A	PPB		08 Mar 2018 14:35:07	U	SMPL
J1810-05ADLX50 - 1		22411	6.1606	-7470A	PPB		08 Mar 2018 14:37:16	U	SMPL
CCV17 - 1	CCV17	18293	5.0310	-7470A	PPB		08 Mar 2018 14:39:23	U	SMPL
CCB17 - 1	CCB17	-238	-0.0521	-7470A	PPB		08 Mar 2018 14:41:33	U	SMPL