



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93764

Temp.: 19.1 °C

SLOPE.: 97.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/08/2018	13:20
2	CAL2	1	Water			7.00	03/08/2018	13:21
3	CAL3	1	Water			10.02	03/08/2018	13:22
4	ICV	1	Water			7.01	03/08/2018	13:24
5	CCV1	1	Water			2.01	03/08/2018	13:25
6	J1850-01	1	Water			7.13	03/08/2018	13:26
7	J1850-01DUP	1	Water			7.14	03/08/2018	13:27
8	J1851-01	1	Water			4.72	03/08/2018	13:30
9	J1852-01	1	Water			4.99	03/08/2018	13:31
10	CCV2	1	Water			12.01	03/08/2018	13:32

1303764

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1853

WorkList ID : 109380

Date : 3/8/2018 1:05:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2018	Water	J1850-01-1	pH	Cool 4 deg C	STER14	H32	JS-SYSTEM	03/05/2018	SM 4500-PH B
03/05/2018	Water	J1851-01-1	pH	Cool 4 deg C	STER14	H42	MINI-CELL	03/05/2018	SM 4500-PH B
03/07/2018	Water	J1852-01-1	pH	Cool 4 deg C	STER05	H43	MINI-CELL	03/07/2018	SM 4500-PH B

Date/Time 03-08-18 11:11 AM
 Received by: JP
 Relinquished by: JP

Date/Time 03-08-18 2:25 PM
 Received by: JP
 Relinquished by: JP