

7471B


LB 93775

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46072



A= 0.0000e+000

B= 2.1826e-004

C= 5.7860e-002

Rho= 0.9996601

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	-272	0.000	-272				
0.2	0.200	0.222	0.022	751	0.0 %	751				
2.5	2.500	2.609	0.109	11689	0.0 %	11689				
5.0	5.000	4.858	-0.142	21995	0.0 %	21995				
7.5	7.500	7.399	-0.101	33634	0.0 %	33634				
10	10.000	10.113	0.113	46072	0.0 %	46072				

LB93775 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 08 Mar 2018 17:15:52				
Sample ID	Extended ID	μ Abs	Conc.	Std Cond	Method	Units	Date	Type
0.0 - 1	50	-272	-	0.0000	7471B	PPB	08 Mar 2018 17:33:35	S Std
0.2 - 1	50.2	751	-	0.2000	7471B	PPB	08 Mar 2018 17:35:52	S Std
2.5 - 1	52.5	11689	-	2.5000	7471B	PPB	08 Mar 2018 17:38:10	S Std
5.0 - 1	55	21995	-	5.0000	7471B	PPB	08 Mar 2018 17:40:27	S Std
7.5 - 1	57.5	33634	-	7.5000	7471B	PPB	08 Mar 2018 17:42:45	S Std
10 - 1	59	46072	-	10.0000	7471B	PPB	08 Mar 2018 17:47:40	S Std
ICV34 - 1	ICV34	17085	3.7868	-	7471B	PPB	08 Mar 2018 17:52:10	U SMPL
ICB34 - 1	ICB34	-196	0.0151	-	7471B	PPB	08 Mar 2018 17:54:26	U SMPL
CCV17 - 1	CCV17	22780	5.0298	-	7471B	PPB	08 Mar 2018 17:56:42	U SMPL
CCB17 - 1	CCB17	-189	0.0166	-	7471B	PPB	08 Mar 2018 17:58:57	U SMPL
CRA - 1	CRA	782	0.2285	-	7471B	PPB	08 Mar 2018 18:01:14	U SMPL
HighStd - 1	HighStd	46669	10.2438	-	7471B	PPB	08 Mar 2018 18:03:30	U SMPL
ChkStd - 1	ChkStd	29139	6.4177	-	7471B	PPB	08 Mar 2018 18:05:46	U SMPL
PB107141BL - 1	PBS	-291	-0.0057	-	7471B	PPB	08 Mar 2018 18:08:04	U SMPL
PB107141BS - 1	LCSS	16288	3.6128	-	7471B	PPB	08 Mar 2018 18:10:22	U SMPL
J1748-01 - 1	AU-06030618-A	1294	0.3403	-	7471B	PPB	08 Mar 2018 18:17:31	U SMPL
J1748-01DUP - 1	AU-06030618-ADUP	1220	0.3241	-	7471B	PPB	08 Mar 2018 18:19:47	U SMPL
J1748-01MS - 1	AU-06030618-AMS	17850	3.9538	-	7471B	PPB	08 Mar 2018 18:22:05	U SMPL
J1748-01MSD - 1	AU-06030618-AMSD	17851	3.9540	-	7471B	PPB	08 Mar 2018 18:24:23	U SMPL
J1748-03 - 1	AU-06030618-B	1016	0.2796	-	7471B	PPB	08 Mar 2018 18:26:39	U SMPL
CCV18 - 1	CCV18	22979	5.0732	-	7471B	PPB	08 Mar 2018 18:28:55	U SMPL
CCB18 - 1	CCB18	-133	0.0288	-	7471B	PPB	08 Mar 2018 18:31:11	U SMPL
J1748-05 - 1	AU-06030618-C	1750	0.4398	-	7471B	PPB	08 Mar 2018 18:33:28	U SMPL
J1749-01 - 1	BU-01-030618	1283	0.3379	-	7471B	PPB	08 Mar 2018 18:35:44	U SMPL
J1760-01 - 1	SU-04-030618	1405	0.3645	-	7471B	PPB	08 Mar 2018 18:38:00	U SMPL
J1811-01 - 1	T-1_030518	7216	1.6328	-	7471B	PPB	08 Mar 2018 18:40:16	U SMPL
J1811-02 - 1	T-2_030518	3995	0.9298	-	7471B	PPB	08 Mar 2018 18:42:33	U SMPL
J1811-03 - 1	T-3_030518	6086	1.3862	-	7471B	PPB	08 Mar 2018 18:44:50	U SMPL
J1811-04 - 1	T-4_030518	5596	1.2792	-	7471B	PPB	08 Mar 2018 18:47:07	U SMPL
J1811-05 - 1	T-5_030518	12104	2.6997	-	7471B	PPB	08 Mar 2018 18:49:25	U SMPL
J1811-06 - 1	T-6_030518	4259	0.9874	-	7471B	PPB	08 Mar 2018 18:51:42	U SMPL
J1811-07 - 1	T-7_030518	4213	0.9774	-	7471B	PPB	08 Mar 2018 18:53:59	U SMPL
CCV19 - 1	CCV19	23152	5.1110	-	7471B	PPB	08 Mar 2018 18:56:15	U SMPL
CCB19 - 1	CCB19	27	0.0638	-	7471B	PPB	08 Mar 2018 18:58:31	U SMPL
J1811-08 - 1	T-8_030518	5568	1.2731	-	7471B	PPB	08 Mar 2018 19:00:48	U SMPL
J1811-09 - 1	T-9_030518	9668	2.1680	-	7471B	PPB	08 Mar 2018 19:03:05	U SMPL
J1811-10 - 1	T-10_030518	4770	1.0990	-	7471B	PPB	08 Mar 2018 19:05:21	U SMPL
J1811-11 - 1	T-11_030518	22263	4.9169	-	7471B	PPB	08 Mar 2018 19:07:37	U SMPL
J1811-12 - 1	T-12_030518	8425	1.8967	-	7471B	PPB	08 Mar 2018 19:09:54	U SMPL
J1811-13 - 1	COMP-1	5986	1.3644	-	7471B	PPB	08 Mar 2018 19:12:12	U SMPL
J1811-14 - 1	COMP-2	6122	1.3940	-	7471B	PPB	08 Mar 2018 19:14:29	U SMPL
J1820-01 - 1	TP-1	247422	54.0597	-	7471B	PPB	08 Mar 2018 19:16:47	U SMPL
PB107142BL - 1	PBS	-131	0.0293	-	7471B	PPB	08 Mar 2018 19:21:55	U SMPL
PB107142BS - 1	LCSS	14920	3.3143	-	7471B	PPB	08 Mar 2018 19:24:11	U SMPL
CCV20 - 1	CCV20	22712	5.0149	-	7471B	PPB	08 Mar 2018 19:26:28	U SMPL
CCB20 - 1	CCB20	-7	0.0563	-	7471B	PPB	08 Mar 2018 19:28:44	U SMPL
J1819-01 - 1	TP-9	18132	4.0153	-	7471B	PPB	08 Mar 2018 19:31:01	U SMPL
J1819-02 - 1	TP-10	731	0.2174	-	7471B	PPB	08 Mar 2018 19:33:18	U SMPL
J1819-02DUP - 1	TP-10DUP	649	0.1995	-	7471B	PPB	08 Mar 2018 19:35:34	U SMPL
J1819-02MS - 1	TP-10MS	16950	3.7573	-	7471B	PPB	08 Mar 2018 19:37:51	U SMPL
J1819-02MSD - 1	TP-10MSD	17048	3.7787	-	7471B	PPB	08 Mar 2018 19:40:07	U SMPL
J1819-03 - 1	TP-11	16975	3.7628	-	7471B	PPB	08 Mar 2018 19:42:24	U SMPL
J1821-01 - 1	COMPOSITE-1-3-20180204	30900	6.8020	-	7471B	PPB	08 Mar 2018 19:44:41	U SMPL
J1821-02 - 1	COMPOSITE-4-5-20180205	24654	5.4388	-	7471B	PPB	08 Mar 2018 19:46:59	U SMPL
J1821-03 - 1	AMENDED-COMPOSITE-1-3-2018	28848	6.3542	-	7471B	PPB	08 Mar 2018 19:49:18	U SMPL
J1821-04 - 1	AMENDED-COMPOSITE-4-5-2018	28003	6.1697	-	7471B	PPB	08 Mar 2018 19:51:35	U SMPL
CCV21 - 1	CCV21	22063	4.8733	-	7471B	PPB	08 Mar 2018 19:53:53	U SMPL
CCB21 - 1	CCB21	-186	0.0173	-	7471B	PPB	08 Mar 2018 19:56:10	U SMPL
J1823-01 - 1	MHB-WCA	9479	2.1267	-	7471B	PPB	08 Mar 2018 19:58:28	U SMPL
J1823-03 - 1	MHB-WCB	627	0.1947	-	7471B	PPB	08 Mar 2018 20:00:44	U SMPL
J1824-01 - 1	FK-BPM-01-E0-E0a-F0-F0a	1476	0.3800	-	7471B	PPB	08 Mar 2018 20:03:00	U SMPL
J1825-01 - 1	FK-BPM-01-C0-C0a-D0-D0a	1357	0.3540	-	7471B	PPB	08 Mar 2018 20:05:16	U SMPL
J1826-08 - 1	TP-2-COMP	135	0.0873	-	7471B	PPB	08 Mar 2018 20:07:33	U SMPL
J1826-09 - 1	TP-3-A-COMP	217	0.1052	-	7471B	PPB	08 Mar 2018 20:09:49	U SMPL
J1826-10 - 1	TP-3-B-COMP	66	0.0723	-	7471B	PPB	08 Mar 2018 20:12:06	U SMPL
J1828-01 - 1	STONE-SUBSTATION	1189	0.3174	-	7471B	PPB	08 Mar 2018 20:14:24	U SMPL
J1828-07 - 1	Concrete-Structure	173	0.0956	-	7471B	PPB	08 Mar 2018 20:16:42	U SMPL
J1829-02 - 1	SB-05-6.5-10.5	19865	4.3936	-	7471B	PPB	08 Mar 2018 20:18:59	U SMPL
CCV22 - 1	CCV22	22039	4.8681	-	7471B	PPB	08 Mar 2018 20:21:16	U SMPL
CCB22 - 1	CCB22	-115	0.0328	-	7471B	PPB	08 Mar 2018 20:23:33	U SMPL
J1829-05 - 1	SB-12-01-05	14299	3.1787	-	7471B	PPB	08 Mar 2018 20:25:50	U SMPL
J1831-01 - 1	1-1-4-BLUE-STONE-DGA	-101	0.0358	-	7471B	PPB	08 Mar 2018 20:28:07	U SMPL
J1748-01LX5 - 1		211	0.1039	-	7471B	PPB	08 Mar 2018 20:30:24	U SMPL
J1748-01A - 1		17528	3.8835	-	7471B	PPB	08 Mar 2018 20:36:53	U SMPL
J1819-02LX5 - 1		-34	0.0504	-	7471B	PPB	08 Mar 2018 20:39:09	U SMPL
J1819-02A - 1		17807	3.9444	-	7471B	PPB	08 Mar 2018 20:41:27	U SMPL
J1820-01DLX10 - 1		25733	5.6743	-	7471B	PPB	08 Mar 2018 20:43:44	U SMPL
CCV23 - 1	CCV23	22901	5.0562	-	7471B	PPB	08 Mar 2018 20:46:02	U SMPL
CCB23 - 1	CCB23	-105	0.0349	-	7471B	PPB	08 Mar 2018 20:48:20	U SMPL