



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : _____

Run Number: LB93791

Temp.: 12.3 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/09/2018	11:35
2	CAL2	1	Water			7.00	03/09/2018	11:36
3	CAL3	1	Water			10.02	03/09/2018	11:37
4	ICV	1	Water			6.99	03/09/2018	11:39
5	CCV1	1	Water			2.01	03/09/2018	11:40
6	J1839-01	1	Water			5.77	03/09/2018	11:41
7	J1839-01DUP	1	Water			5.78	03/09/2018	11:42
8	J1839-02	1	Water			6.35	03/09/2018	11:44
9	J1839-03	1	Water			6.35	03/09/2018	11:45
10	J1839-04	1	Water			6.56	03/09/2018	11:46
11	CCV2	1	Water			12.02	03/09/2018	11:48

7393791

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1839

WorkList ID : 109435

Date : 3/9/2018 11:28:23 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/06/2018	Water	J1839-01	pH	Cool 4 deg C	LOCK01	H62	MW-1	03/06/2018	9045C
03/06/2018	Water	J1839-02	pH	Cool 4 deg C	LOCK01	H62	MW-3	03/06/2018	9045C
03/06/2018	Water	J1839-03	pH	Cool 4 deg C	LOCK01	H62	MW-4	03/06/2018	9045C
03/06/2018	Water	J1839-04	pH	Cool 4 deg C	LOCK01	H62	MW-2	03/06/2018	9045C

Date/Time 03-09-18 11:30 AM
 Received by: JL
 Relinquished by: LG

Date/Time 03-09-18 11:00 PM
 Received by: JL
 Relinquished by: JL