



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : sweta

Run Number: LB93820

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/09/2018	16:35
2	CAL2	1	Water			7.00	03/09/2018	16:36
3	CAL3	1	Water			10.02	03/09/2018	16:37
4	ICV	1	Water			7.01	03/09/2018	16:39
5	CCV1	1	Water			2.01	03/09/2018	16:40
6	J1776-07	1	Solid	20.02	20	6.55	03/09/2018	16:41
7	J1776-07DUP	1	Solid	20.04	20	6.56	03/09/2018	16:42
8	J1776-13	1	Solid	20.05	20	7.20	03/09/2018	16:43
9	J1776-19	1	Solid	20.03	20	6.09	03/09/2018	16:45
10	J1776-25	1	Solid	20.01	20	5.69	03/09/2018	16:46
11	J1870-03	1	Solid	20.04	20	8.92	03/09/2018	16:47
12	J1873-04	1	Solid	20.06	20	6.75	03/09/2018	16:48
13	J1873-07	1	Solid	20.04	20	5.59	03/09/2018	16:50
14	J1873-10	1	Solid	20.02	20	8.34	03/09/2018	16:52
15	J1874-02	1	Solid	20.03	20	8.57	03/09/2018	17:00
16	CCV2	1	Water			12.02	03/09/2018	17:01
17	J1876-03	1	Solid	20.04	20	11.22	03/09/2018	17:03
18	J1876-06	1	Solid	20.02	20	11.07	03/09/2018	17:05
19	CCV3	1	Water			2.01	03/09/2018	17:06

18 93820

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1876 WorkList ID : 109455 Date : 3/9/2018 4:24:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/09/2018	Solid	J1776-07	Corrosivity	Cool 4 deg C	PSEG01	I51	TP-1-3FT	03/09/2018	9045C
03/09/2018	Solid	J1776-13	Corrosivity	Cool 4 deg C	PSEG01	I51	TP-1-6FT	03/09/2018	9045C
03/09/2018	Solid	J1776-19	Corrosivity	Cool 4 deg C	PSEG01	I51	TP-2-3FT	03/09/2018	9045C
03/09/2018	Solid	J1776-25	Corrosivity	Cool 4 deg C	PSEG01	I51	TP-2-6FT	03/09/2018	9045C
03/09/2018	Solid	J1870-03	Corrosivity	Cool 4 deg C	GZAG04	H31	WC-06-08-12	03/09/2018	9045C
03/09/2018	Solid	J1873-04	Corrosivity	Cool 4 deg C	PSEG01	I51	TP-8-COMP	03/09/2018	9045C
03/09/2018	Solid	J1873-07	Corrosivity	Cool 4 deg C	PSEG01	I51	TP-9A-COMP	03/09/2018	9045C
03/09/2018	Solid	J1873-10	Corrosivity	Cool 4 deg C	PSEG01	I51	TP-9B-COMP	03/09/2018	9045C
03/09/2018	Solid	J1874-02	Corrosivity	Cool 4 deg C	PSEG01	I51	MHC-WCIS	03/09/2018	9045C
03/08/2018	Solid	J1876-03	Corrosivity	Cool 4 deg C	GZAG04	H61	WC-04-04-10	03/08/2018	9045C
03/08/2018	Solid	J1876-06	Corrosivity	Cool 4 deg C	GZAG04	H61	WC-05-01-03	03/08/2018	9045C

Date/Time 03.09.18 9.30 PM
Received by: JS
Relinquished by: JS

Date/Time 03.09.18 6.35 PM
Received by: cy
Relinquished by: JS