


7471B

LB93841
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42475



A= 0.0000e+000

B= 2.3205e-004

C= 4.9767e-002

Rho= 0.9996642

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.040	0.040	-41	0.000	-41				
0.2	0.200	0.202	0.002	654	0.0 %	654				
2.5	2.500	2.439	-0.061	10294	0.0 %	10294				
5.0	5.000	4.926	-0.074	21013	0.0 %	21013				
7.5	7.500	7.688	0.188	32914	0.0 %	32914				
10	10.000	9.906	-0.094	42475	0.0 %	42475				

LB93841 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Mar 2018 16:09:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	7471B	PPB	12 Mar 2018 16:11:58	S	Std
0.2 - 1	50.2	654	-	0.2000	7471B	PPB	12 Mar 2018 16:14:15	S	Std
2.5 - 1	52.5	10294	-	2.5000	7471B	PPB	12 Mar 2018 16:16:33	S	Std
5.0 - 1	55	21013	-	5.0000	7471B	PPB	12 Mar 2018 16:18:50	S	Std
7.5 - 1	57.5	32914	-	7.5000	7471B	PPB	12 Mar 2018 16:21:08	S	Std
10 - 1	510	42475	-	10.0000	7471B	PPB	12 Mar 2018 16:26:17	S	Std
ICV39 - 1	ICV39	16305	3.8334	-	7471B	PPB	12 Mar 2018 16:29:03	U	SMPL
ICB39 - 1	ICB39	-222	-0.0017	-	7471B	PPB	12 Mar 2018 16:31:18	U	SMPL
CCV39 - 1	CCV39	20990	4.9206	-	7471B	PPB	12 Mar 2018 16:33:34	U	SMPL
CCB39 - 1	CCB39	-211	0.0008	-	7471B	PPB	12 Mar 2018 16:35:49	U	SMPL
CRA - 1	CRA	601	0.1892	-	7471B	PPB	12 Mar 2018 16:38:06	U	SMPL
HighStd - 1	HighStd	42509	9.9141	-	7471B	PPB	12 Mar 2018 16:40:22	U	SMPL
ChkStd - 1	ChkStd	28283	6.6129	-	7471B	PPB	12 Mar 2018 16:42:38	U	SMPL
PB107228BL - 1	PBS	-217	-0.0006	-	7471B	PPB	12 Mar 2018 16:44:56	U	SMPL
PB107228BS - 1	LCSS	15078	3.5487	-	7471B	PPB	12 Mar 2018 16:47:15	U	SMPL
J1776-02 - 1	TP-1-3FT	2153	0.5494	-	7471B	PPB	12 Mar 2018 16:49:31	U	SMPL
J1776-08 - 1	TP-1-6FT	1283	0.3475	-	7471B	PPB	12 Mar 2018 16:51:48	U	SMPL
J1776-14 - 1	TP-2-3FT	968	0.2744	-	7471B	PPB	12 Mar 2018 16:54:06	U	SMPL
J1776-20 - 1	TP-2-6FT	1770	0.4605	-	7471B	PPB	12 Mar 2018 16:56:23	U	SMPL
J1870-02 - 1	SB-10-08-12	37222	8.6873	-	7471B	PPB	12 Mar 2018 16:58:40	U	SMPL
CCV40 - 1	CCV40	20654	4.8426	-	7471B	PPB	12 Mar 2018 17:00:56	U	SMPL
CCB40 - 1	CCB40	-220	-0.0013	-	7471B	PPB	12 Mar 2018 17:03:13	U	SMPL
J1873-03 - 1	TP-8-COMP	1348	0.3626	-	7471B	PPB	12 Mar 2018 17:05:29	U	SMPL
J1873-06 - 1	TP-9A-COMP	338	0.1282	-	7471B	PPB	12 Mar 2018 17:07:45	U	SMPL
J1873-09 - 1	TP-9B-COMP	199	0.0959	-	7471B	PPB	12 Mar 2018 17:10:00	U	SMPL
J1874-01 - 1	MHC-WCIS	4950	1.1984	-	7471B	PPB	12 Mar 2018 17:12:17	U	SMPL
J1876-02 - 1	SB-07-08-10	34161	7.9769	-	7471B	PPB	12 Mar 2018 17:14:33	U	SMPL
J1876-04 - 1	SB-11-01-03	5930	1.4258	-	7471B	PPB	12 Mar 2018 17:16:50	U	SMPL
J1880-01 - 1	SAMPLE-F	2721	0.6812	-	7471B	PPB	12 Mar 2018 17:19:08	U	SMPL
J1880-01DUP - 1	SAMPLE-FDUP	2469	0.6227	-	7471B	PPB	12 Mar 2018 17:21:26	U	SMPL
J1880-01MS - 1	SAMPLE-FMS	18721	4.3940	-	7471B	PPB	12 Mar 2018 17:23:44	U	SMPL
J1880-01MSD - 1	SAMPLE-FMSD	18682	4.3850	-	7471B	PPB	12 Mar 2018 17:26:00	U	SMPL
CCV41 - 1	CCV41	20976	4.9173	-	7471B	PPB	12 Mar 2018 17:28:16	U	SMPL
CCB41 - 1	CCB41	-195	0.0045	-	7471B	PPB	12 Mar 2018 17:30:33	U	SMPL
J1880-01LX5 - 1		435	0.1507	-	7471B	PPB	12 Mar 2018 17:32:49	U	SMPL
J1880-01A - 1		19641	4.6075	-	7471B	PPB	12 Mar 2018 17:35:05	U	SMPL
CCV42 - 1	CCV42	20994	4.9215	-	7471B	PPB	12 Mar 2018 17:37:22	U	SMPL
CCB42 - 1	CCB42	-195	0.0045	-	7471B	PPB	12 Mar 2018 17:39:38	U	SMPL