



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93843

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/12/2018	16:30
2	CAL2	1	Water			7.01	03/12/2018	16:31
3	CAL3	1	Water			10.02	03/12/2018	16:34
4	ICV	1	Water			6.99	03/12/2018	16:35
5	CCV1	1	Water			2.01	03/12/2018	16:36
6	J1753-02	1	Solid	20.02	20	8.52	03/12/2018	16:37
7	J1753-02DUP	1	Solid	20.04	20	8.53	03/12/2018	16:38
8	J1890-02	1	Solid	20.03	20	8.61	03/12/2018	16:40
9	J1893-02	1	Solid	20.05	20	7.25	03/12/2018	16:41
10	J1898-04	1	Solid	20.01	20	5.85	03/12/2018	16:42
11	J1898-08	1	Solid	20.03	20	7.85	03/12/2018	16:44
12	J1898-10	1	Solid	20.02	20	7.75	03/12/2018	16:45
13	J1898-14	1	Solid	20.04	20	6.72	03/12/2018	16:46
14	CCV2	1	Water			12.01	03/12/2018	16:47

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1890

WorkList ID : 109513

Date : 3/12/2018 1:44:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2018	Solid	J1753-02	Corrosivity	Cool 4 deg C	PSEG05	I53	HD-02-031218	03/12/2018	9045C
03/12/2018	Solid	J1890-02	Corrosivity	Cool 4 deg C	PSEG01	I51	DRUM	03/12/2018	9045C
03/12/2018	Solid	J1893-02	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-1	03/12/2018	9045C
03/12/2018	Solid	J1898-04	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-10-COMP	03/12/2018	9045C
03/12/2018	Solid	J1898-08	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-11A-COMP	03/12/2018	9045C
03/12/2018	Solid	J1898-10	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-11B-COMP	03/12/2018	9045C
03/12/2018	Solid	J1898-14	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-12-COMP	03/12/2018	9045C

Date/Time 03-12-18 4:10 PM
 Received by: JD
 Relinquished by: JD

Date/Time 03-12-18 6:00 PM
 Received by: JD
 Relinquished by: JD