



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93844

Temp.: 18.2 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/12/2018	17:10
2	CAL2	1	Water			6.99	03/12/2018	17:11
3	CAL3	1	Water			10.04	03/12/2018	17:12
4	ICV	1	Water			7.00	03/12/2018	17:13
5	CCV1	1	Water			2.01	03/12/2018	17:15
6	J1898-16	1	Water			7.06	03/12/2018	17:16
7	J1898-16DUP	1	Water			7.07	03/12/2018	17:17
8	CCV2	1	Water			12.02	03/12/2018	17:18

WORKLIST(Hardcopy Internal Chain)

23 93844

WorkList Name : corrosivity j1898 WorkList ID : 109518 Date : 3/12/2018 4:45:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2018	Water	J1898-16-6	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-11	03/12/2018	9040C

Date/Time 03-12-18 4:50 PM
 Received by: JP
 Relinquished by: JP

Date/Time 03-12-18 6:10 PM
 Received by: JP
 Relinquished by: JP