

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/13/2018

DATE & TIME IN: 03/13/2018 09:30

OVEN TEMP: 103 °C

Run Number: LB93852

DATE & TIME OUT: 03/13/2018 14:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB93852BL	LB93852BL	1.7846	1.7846	100	1.7846	1.7846	1.7846	0	0
2	LB93852BS	LB93852BS	1.8163	1.8163	100	1.867	1.867	1.867	0.0507	507
3	J1889-01	EFFWW	1.8227	1.8227	100	1.8367	1.8367	1.8367	0.014	140
4	J1894-01	031218-TIDO-F-D-B	1.8317	1.8317	300	1.8404	1.8404	1.8404	0.0087	29
5	J1894-02	031218-TIDO-F-D-S	1.824	1.824	300	1.8328	1.8328	1.8328	0.0088	29.3
6	J1894-03	031218-TIDO-F-U-B	1.8313	1.8313	300	1.8442	1.8442	1.8442	0.0129	43
7	J1894-04	031218-TIDO-F-U-S	1.8308	1.8308	300	1.8404	1.8404	1.8404	0.0096	32
8	J1895-01	030918-TIDO-E-U-B	1.8341	1.8341	200	1.8414	1.8414	1.8414	0.0073	36.5
9	J1895-02	030918-TIDO-E-U-S	1.8095	1.8095	300	1.8214	1.8214	1.8214	0.0119	39.7
10	J1895-03	030918-TIDO-E-D-B	1.8142	1.8142	200	1.8212	1.8212	1.8212	0.007	35
11	J1895-04	030918-TIDO-E-D-S	1.8195	1.8195	200	1.8259	1.8259	1.8259	0.0064	32
12	J1898-15	TP-11	1.8096	1.8096	100	1.8302	1.8302	1.8302	0.0206	206
13	J1898-15DUP	TP-11DUP	1.831	1.831	100	1.8513	1.8513	1.8513	0.0203	203

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS031318

WorkList ID : 109537

Date : 3/13/2018 8:37:13 AM

2393852

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2018	Water	J1889-01 C	TSS	Cool 4 deg C	ARDM01	H63	EFFWW	03/12/2018	SM2540 D
03/18/2018	Water	J1894-01 C	TSS	Cool 4 deg C	HDR103	H62	031218-TIDO-F-D-B	03/12/2018	SM2540 D
03/18/2018	Water	J1894-02 C	TSS	Cool 4 deg C	HDR103	H62	031218-TIDO-F-D-S	03/12/2018	SM2540 D
03/18/2018	Water	J1894-03 C	TSS	Cool 4 deg C	HDR103	H62	031218-TIDO-F-U-B	03/12/2018	SM2540 D
03/18/2018	Water	J1894-04 C	TSS	Cool 4 deg C	HDR103	H62	031218-TIDO-F-U-S	03/12/2018	SM2540 D
03/16/2018	Water	J1895-01 C	TSS	Cool 4 deg C	HDR103	H11	030918-TIDO-E-U-B	03/09/2018	SM2540 D
03/16/2018	Water	J1895-02 C	TSS	Cool 4 deg C	HDR103	H11	030918-TIDO-E-U-S	03/09/2018	SM2540 D
03/16/2018	Water	J1895-03 C	TSS	Cool 4 deg C	HDR103	H11	030918-TIDO-E-D-B	03/09/2018	SM2540 D
03/16/2018	Water	J1895-04 C	TSS	Cool 4 deg C	HDR103	H11	030918-TIDO-E-D-S	03/09/2018	SM2540 D
03/15/2018	Water	J1898-15 F	TSS	Cool 4 deg C	PSEG01	I53	TP-11	03/12/2018	SM2540 D

Date/Time 03/13/18 08:50
 Received by: AS
 Relinquished by: DP

Date/Time 03/13/18, 9:40
 Received by: DP
 Relinquished by: AS