


7470A

LB 93854
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35349



A= 0.0000e+000

B= 2.8357e-004

C= 7.2023e-002

Rho= 0.9996206

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.058	0.058	-49	0.000	-49				
0.2	0.200	0.227	0.027	546	0.0 %	546				
2.5	2.500	2.542	0.042	8711	0.0 %	8711				
5.0	5.000	4.783	-0.217	16615	0.0 %	16615				
7.5	7.500	7.494	-0.006	26172	0.0 %	26172				
10.0	10.000	10.096	0.096	35349	0.0 %	35349				

LB93854 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Mar 2018 10:09:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	-49	-	0.00007470A	PPB		13 Mar 2018 10:24:30	S	Std
0.2 - 1	50.2	546	-	0.20007470A	PPB		13 Mar 2018 10:26:38	S	Std
2.5 - 1	52.5	8711	-	2.50007470A	PPB		13 Mar 2018 10:28:46	S	Std
5.0 - 1	55	16615	-	5.00007470A	PPB		13 Mar 2018 10:30:54	S	Std
7.5 - 1	57.5	26172	-	7.50007470A	PPB		13 Mar 2018 10:33:02	S	Std
10.0 - 1	59	35349	-	10.00007470A	PPB		13 Mar 2018 10:35:10	S	Std
ICV40 - 1	ICV40	13830	3.9938	-7470A	PPB		13 Mar 2018 10:38:00	U	SMPL
ICB40 - 1	ICB40	43	0.0842	-7470A	PPB		13 Mar 2018 10:40:07	U	SMPL
CCV43 - 1	CCV43	17634	5.0724	-7470A	PPB		13 Mar 2018 10:44:45	U	SMPL
CCB43 - 1	CCB43	-85	0.0479	-7470A	PPB		13 Mar 2018 10:46:51	U	SMPL
CRA - 1	CRA	443	0.1976	-7470A	PPB		13 Mar 2018 10:51:25	U	SMPL
HighStd - 1	HighStd	36165	10.3272	-7470A	PPB		13 Mar 2018 10:53:31	U	SMPL
ChkStd - 1	ChkStd	24689	7.0730	-7470A	PPB		13 Mar 2018 10:55:38	U	SMPL
PB107255BL - 1	PBW	-266	-0.0034	-7470A	PPB		13 Mar 2018 10:57:47	U	SMPL
PB107255BS - 1	LCSW	13548	3.9138	-7470A	PPB		13 Mar 2018 10:59:56	U	SMPL
J1867-01 - 1	CPS133071	274	0.1497	-7470A	PPB		13 Mar 2018 11:02:04	U	SMPL
J1867-02 - 1	CPS113071	-187	0.0190	-7470A	PPB		13 Mar 2018 11:04:12	U	SMPL
J1867-03 - 1	CPS093071	-164	0.0255	-7470A	PPB		13 Mar 2018 11:06:21	U	SMPL
J1867-04 - 1	CPS043071	-165	0.0252	-7470A	PPB		13 Mar 2018 11:08:29	U	SMPL
J1867-05 - 1	CPS043075	-176	0.0221	-7470A	PPB		13 Mar 2018 11:10:36	U	SMPL
CCV44 - 1	CCV44	17211	4.9525	-7470A	PPB		13 Mar 2018 11:12:43	U	SMPL
CCB44 - 1	CCB44	-262	-0.0023	-7470A	PPB		13 Mar 2018 11:14:50	U	SMPL
J1867-06 - 1	CPS123071	43	0.0842	-7470A	PPB		13 Mar 2018 11:16:57	U	SMPL
J1867-07 - 1	CPS083071	239	0.1398	-7470A	PPB		13 Mar 2018 11:19:04	U	SMPL
J1867-08 - 1	CPS103071	-207	0.0133	-7470A	PPB		13 Mar 2018 11:21:11	U	SMPL
J1867-09 - 1	C19013071	-148	0.0301	-7470A	PPB		13 Mar 2018 11:23:18	U	SMPL
J1867-10 - 1	CPS073071	-154	0.0284	-7470A	PPB		13 Mar 2018 11:25:25	U	SMPL
J1867-11 - 1	C19013072	-178	0.0215	-7470A	PPB		13 Mar 2018 11:27:33	U	SMPL
J1867-12 - 1	C190137	-165	0.0252	-7470A	PPB		13 Mar 2018 11:29:41	U	SMPL
J1867-12DUP - 1	C190137DUP	-155	0.0281	-7470A	PPB		13 Mar 2018 11:31:49	U	SMPL
J1867-13 - 1	J1867-12MS	12807	3.7037	-7470A	PPB		13 Mar 2018 11:33:57	U	SMPL
J1867-14 - 1	J1867-12MSD	12778	3.6954	-7470A	PPB		13 Mar 2018 11:36:04	U	SMPL
CCV45 - 1	CCV45	17646	5.0758	-7470A	PPB		13 Mar 2018 11:38:12	U	SMPL
CCB45 - 1	CCB45	-239	0.0043	-7470A	PPB		13 Mar 2018 11:40:20	U	SMPL
PB107256BL - 1	PBW	-98	0.0442	-7470A	PPB		13 Mar 2018 11:42:27	U	SMPL
PB107256BS - 1	LCSW	12797	3.7008	-7470A	PPB		13 Mar 2018 11:44:34	U	SMPL
J1867-15 - 1	CPS133071	-23	0.0655	-7470A	PPB		13 Mar 2018 11:46:41	U	SMPL
J1867-16 - 1	CPS113071	-214	0.0113	-7470A	PPB		13 Mar 2018 11:48:48	U	SMPL
J1867-17 - 1	CPS093071	-179	0.0213	-7470A	PPB		13 Mar 2018 11:50:56	U	SMPL
J1867-18 - 1	CPS043071	-192	0.0176	-7470A	PPB		13 Mar 2018 11:53:03	U	SMPL
J1867-19 - 1	CPS123071	-121	0.0377	-7470A	PPB		13 Mar 2018 11:55:12	U	SMPL
J1867-21 - 1	CPS083071	-139	0.0326	-7470A	PPB		13 Mar 2018 11:57:20	U	SMPL
J1867-22 - 1	CPS103071	-156	0.0278	-7470A	PPB		13 Mar 2018 11:59:28	U	SMPL
J1867-23 - 1	C19013071	-126	0.0363	-7470A	PPB		13 Mar 2018 12:01:35	U	SMPL
CCV46 - 1	CCV46	17512	5.0378	-7470A	PPB		13 Mar 2018 12:03:42	U	SMPL
CCB46 - 1	CCB46	-214	0.0113	-7470A	PPB		13 Mar 2018 12:05:49	U	SMPL
J1867-24 - 1	CPS073071	-173	0.0230	-7470A	PPB		13 Mar 2018 12:07:57	U	SMPL
J1867-25 - 1	C19013072	-126	0.0363	-7470A	PPB		13 Mar 2018 12:10:05	U	SMPL
J1867-26 - 1	C190137	-117	0.0388	-7470A	PPB		13 Mar 2018 12:12:11	U	SMPL
J1867-26DUP - 1	C190137DUP	-160	0.0267	-7470A	PPB		13 Mar 2018 12:14:18	U	SMPL
J1867-27 - 1	J1867-26MS	12902	3.7306	-7470A	PPB		13 Mar 2018 12:16:26	U	SMPL
J1867-28 - 1	J1867-26MSD	12648	3.6586	-7470A	PPB		13 Mar 2018 12:18:33	U	SMPL
PB107257BL - 1	PBW	-279	-0.0071	-7470A	PPB		13 Mar 2018 12:20:42	U	SMPL
PB107257BS - 1	LCSW	12857	3.7178	-7470A	PPB		13 Mar 2018 12:22:51	U	SMPL
PB107257TB - 1	PB107184TB	885	0.3230	-7470A	PPB		13 Mar 2018 12:25:00	U	SMPL
J1821-08 - 1	AMENDED-COMPOSITE-1-3-20180204	770	0.2904	-7470A	PPB		13 Mar 2018 12:27:07	U	SMPL
CCV47 - 1	CCV47	16520	4.7565	-7470A	PPB		13 Mar 2018 12:29:14	U	SMPL
CCB47 - 1	CCB47	-196	0.0164	-7470A	PPB		13 Mar 2018 12:31:21	U	SMPL
J1821-09 - 1	AMENDED-COMPOSITE-4-5-20180205	349	0.1710	-7470A	PPB		13 Mar 2018 12:33:29	U	SMPL
J1821-09DUP - 1	AMENDED-COMPOSITE-4-5-20180205DUP	370	0.1769	-7470A	PPB		13 Mar 2018 12:35:36	U	SMPL
J1821-09MS - 1	AMENDED-COMPOSITE-4-5-20180205MS	13587	3.9248	-7470A	PPB		13 Mar 2018 12:37:43	U	SMPL
J1821-09MSD - 1	AMENDED-COMPOSITE-4-5-20180205MSD	12801	3.7020	-7470A	PPB		13 Mar 2018 12:39:50	U	SMPL
PB107258BL - 1	PBW	-217	0.0105	-7470A	PPB		13 Mar 2018 12:41:58	U	SMPL
PB107258BS - 1	LCSW	12773	3.6940	-7470A	PPB		13 Mar 2018 12:44:06	U	SMPL
PB107258TB - 1	PB107226TB	-190	0.0181	-7470A	PPB		13 Mar 2018 12:46:14	U	SMPL
J1776-07 - 1	TP-1-3FT	-98	0.0442	-7470A	PPB		13 Mar 2018 12:48:23	U	SMPL
J1776-13 - 1	TP-1-6FT	-95	0.0451	-7470A	PPB		13 Mar 2018 12:50:32	U	SMPL
J1776-19 - 1	TP-2-3FT	-76	0.0505	-7470A	PPB		13 Mar 2018 12:52:39	U	SMPL
CCV48 - 1	CCV48	17590	5.0600	-7470A	PPB		13 Mar 2018 12:54:47	U	SMPL
CCB48 - 1	CCB48	-206	0.0136	-7470A	PPB		13 Mar 2018 12:56:54	U	SMPL
J1776-25 - 1	TP-2-6FT	-101	0.0434	-7470A	PPB		13 Mar 2018 12:59:02	U	SMPL
J1776-25DUP - 1	TP-2-6FTDUP	-79	0.0496	-7470A	PPB		13 Mar 2018 13:01:10	U	SMPL
J1776-25MS - 1	TP-2-6FTMS	12613	3.6487	-7470A	PPB		13 Mar 2018 13:03:17	U	SMPL
J1776-25MSD - 1	TP-2-6FTMSD	12746	3.6864	-7470A	PPB		13 Mar 2018 13:05:24	U	SMPL
J1873-04 - 1	TP-8-COMP	-215	0.0111	-7470A	PPB		13 Mar 2018 13:07:31	U	SMPL
J1873-07 - 1	TP-9A-COMP	-112	0.0403	-7470A	PPB		13 Mar 2018 13:09:40	U	SMPL
J1873-10 - 1	TP-9B-COMP	-67	0.0530	-7470A	PPB		13 Mar 2018 13:11:48	U	SMPL
J1874-02 - 1	MHC-WCIS	-87	0.0474	-7470A	PPB		13 Mar 2018 13:13:56	U	SMPL
CCV49 - 1	CCV49	17638	5.0736	-7470A	PPB		13 Mar 2018 13:18:12	U	SMPL
CCB49 - 1	CCB49	-193	0.0173	-7470A	PPB		13 Mar 2018 13:20:20	U	SMPL
J1867-12LX5 - 1		-88	0.0471	-7470A	PPB		13 Mar 2018 13:22:28	U	SMPL
J1867-12A - 1		13141	3.7984	-7470A	PPB		13 Mar 2018 13:24:36	U	SMPL
J1867-26LX5 - 1		-176	0.0221	-7470A	PPB		13 Mar 2018 13:26:44	U	SMPL
J1867-26A - 1		13464	3.8900	-7470A	PPB		13 Mar 2018 13:28:51	U	SMPL
J1821-09LX5 - 1		-110	0.0408	-7470A	PPB		13 Mar 2018 13:30:59	U	SMPL
J1821-09A - 1		14019	4.0473	-7470A	PPB		13 Mar 2018 13:33:08	U	SMPL
J1776-25LX5 - 1		-181	0.0207	-7470A	PPB		13 Mar 2018 13:35:15	U	SMPL
J1776-25A - 1		13896	4.0125	-7470A	PPB		13 Mar 2018 13:37:24	U	SMPL
CCV50 - 1	CCV50	17125	4.9281	-7470A	PPB		13 Mar 2018 13:39:32	U	SMPL
CCB50 - 1	CCB50	-198	0.0159	-7470A	PPB		13 Mar 2018 13:41:41	U	SMPL