

2093867

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/13/2018 14:01

Test: Total CN Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.534	0.0	0.098	
ICB1	-0.419	0.0	0.004	
CCV1	244.993	0.0	0.240	
CCB1	-0.499	0.0	0.004	
PB107298BL	-0.489	0.0	0.004	
PB107298BS	97.261	0.0	0.098	
MIDPB107298	246.876	0.0	0.242	
J1889-02	-0.045	0.0	0.004	
J1889-02DUP	-0.123	0.0	0.004	
J1889-02MS	33.447	0.0	0.036	
J1889-02MSD	33.619	0.0	0.036	
J1900-23	302.108	0.0	0.295	
CCV2	257.021	0.0	0.252	
CCB2	-0.481	0.0	0.004	

N 14
Mean 93.629
SD 116.6163
CV% 124.55

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Mar 13 13:35:05 2018

Tue Mar 13 14:00:33 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.6823	µg/l	3/13/2018 11:59:45	
5.0PPBCN	A	Total CN	P	4.2797	µg/l	3/13/2018 11:59:46	
10PPBCN	A	Total CN	P	9.2427	µg/l	3/13/2018 11:59:47	
50PPBCN	A	Total CN	P	50.3072	µg/l	3/13/2018 11:59:48	
100PPBCN	A	Total CN	P	101.3057	µg/l	3/13/2018 11:59:49	
250PPBCN	A	Total CN	P	251.633	µg/l	3/13/2018 11:59:50	
500PPBCN	A	Total CN	P	498.914	µg/l	3/13/2018 11:59:51	
ICV1	S	Total CN	P	97.5344	µg/l	3/13/2018 13:35:05	
ICB1	S	Total CN	P	-0.419	µg/l	3/13/2018 13:35:06	
CCV1	S	Total CN	P	244.9934	µg/l	3/13/2018 13:35:07	
CCB1	S	Total CN	P	-0.4987	µg/l	3/13/2018 13:35:08	
PB107298BL	S	Total CN	P	-0.4886	µg/l	3/13/2018 13:35:09	
PB107298BS	S	Total CN	P	97.2609	µg/l	3/13/2018 13:35:10	
MIDPB107298	S	Total CN	P	246.8755	µg/l	3/13/2018 13:35:11	
J1889-02	S	Total CN	P	-0.0451	µg/l	3/13/2018 13:35:12	
J1889-02DUP	S	Total CN	P	-0.1229	µg/l	3/13/2018 13:35:13	
J1889-02MS	S	Total CN	P	33.4469	µg/l	3/13/2018 13:35:14	
J1889-02MSD	S	Total CN	P	33.6186	µg/l	3/13/2018 13:35:15	
J1900-23	S	Total CN	P	302.1075	µg/l	3/13/2018 13:38:56	
CCV2	S	Total CN	P	257.021	µg/l	3/13/2018 13:38:57	
CCB2	S	Total CN	P	-0.4814	µg/l	3/13/2018 13:38:58	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/13/2018 12:00


Test Total CN

Accepted 3/13/2018 12:00

Factor 1038
Bias 0.004

Coeff. of det. 0.999965

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.6823	0.0000	
2	5.0PPBCN	0.008	4.2797	5.0000	
3	10PPBCN	0.013	9.2427	10.0000	
4	50PPBCN	0.053	50.3072	50.0000	
5	100PPBCN	0.102	101.3057	100.0000	
6	250PPBCN	0.246	251.6330	250.0000	
7	500PPBCN	0.485	498.9140	500.0000	