

LB93879


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47248



A= 0.0000e+000

B= 2.1234e-004

C= -4.5374e-002

Rho= 0.9999619

Accept=Accepted

slope
4.1315e-4

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.00	0.000	-0.031	-0.031	66	0.000	66					-
0.2	0.200	0.192	-0.008	1119	0.0 %	1119					-4
2.5	2.500	2.566	0.066	12296	0.0 %	12296					3
5.0	5.000	4.975	-0.025	23645	0.0 %	23645					-1
7.5	7.500	7.511	0.011	35587	0.0 %	35587					0
10.0	10.000	9.987	-0.013	47248	0.0 %	47248					0

103
3/13/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	66	-		0 ISM02.4	PPB	3/13/2018 16:31	Std
	0.2 S0.2	1119	-		0.2 ISM02.4	PPB	3/13/2018 16:34	Std
	2.5 S2.5	12296	-		2.5 ISM02.4	PPB	3/13/2018 16:36	Std
	5 S5	23645	-		5 ISM02.4	PPB	3/13/2018 16:38	Std
	7.5 S7.5	35587	-		7.5 ISM02.4	PPB	3/13/2018 16:40	Std
	10 S10	47248	-		10 ISM02.4	PPB	3/13/2018 16:43	Std
ICV	ICV	18042	3.7856	-	ISM02.4	PPB	3/13/2018 16:45	SMPL
ICB	ICB	-78	-0.0619	-	ISM02.4	PPB	3/13/2018 16:48	SMPL
CCV054	CCV054	23900	5.0295	-	ISM02.4	PPB	3/13/2018 16:50	SMPL
CCB054	CCB054	-97	-0.066	-	ISM02.4	PPB	3/13/2018 16:52	SMPL
PB107285BL	PBS004	39	-0.0371	-	ISM02.4	PPB	3/13/2018 17:02	SMPL
J1836-01	MF6F00	1163	0.2016	-	ISM02.4	PPB	3/13/2018 17:04	SMPL
J1836-02	MF6F01	1935	0.3655	-	ISM02.4	PPB	3/13/2018 17:07	SMPL
J1836-03	MF6F02	1034	0.1742	-	ISM02.4	PPB	3/13/2018 17:09	SMPL
J1836-04	MF6F03	1089	0.1859	-	ISM02.4	PPB	3/13/2018 17:11	SMPL
J1836-05	MF6F04	1187	0.2067	-	ISM02.4	PPB	3/13/2018 17:13	SMPL
J1836-06	MF6F05	1146	0.198	-	ISM02.4	PPB	3/13/2018 17:16	SMPL
J1836-07	MF6F40	1293	0.2292	-	ISM02.4	PPB	3/13/2018 17:18	SMPL
J1836-08	MF6F41	951	0.1566	-	ISM02.4	PPB	3/13/2018 17:20	SMPL
J1836-09	MF6F42	1014	0.1699	-	ISM02.4	PPB	3/13/2018 17:23	SMPL
J1836-10	MF6F43	938	0.1538	-	ISM02.4	PPB	3/13/2018 17:25	SMPL
J1836-11	MF6F44	1186	0.2065	-	ISM02.4	PPB	3/13/2018 17:27	SMPL
J1836-12	MF6F16	1225	0.2147	-	ISM02.4	PPB	3/13/2018 17:29	SMPL
J1836-13	MF6F17	1108	0.1899	-	ISM02.4	PPB	3/13/2018 17:32	SMPL
J1836-14	MF6F18	1138	0.1963	-	ISM02.4	PPB	3/13/2018 17:34	SMPL
J1836-15	MF6F18D	1089	0.1859	-	ISM02.4	PPB	3/13/2018 17:36	SMPL
J1836-16	MF6F18S	13025	2.7203	-	ISM02.4	PPB	3/13/2018 17:38	SMPL
CCV055	CCV055	23680	4.9828	-	ISM02.4	PPB	3/13/2018 17:41	SMPL
CCB055	CCB055	-83	-0.063	-	ISM02.4	PPB	3/13/2018 17:43	SMPL
J1836-17	MF6F20	1743	0.3247	-	ISM02.4	PPB	3/13/2018 17:45	SMPL
J1836-18	MF6F21	1821	0.3413	-	ISM02.4	PPB	3/13/2018 17:48	SMPL
J1836-19	MF6F24	886	0.1428	-	ISM02.4	PPB	3/13/2018 17:50	SMPL
J1836-20	MF6F28	1267	0.2237	-	ISM02.4	PPB	3/13/2018 17:52	SMPL
J1836-21	MF6F29	4150	0.8358	-	ISM02.4	PPB	3/13/2018 17:54	SMPL
J1836-22	MF6F30	936	0.1534	-	ISM02.4	PPB	3/13/2018 17:57	SMPL
PB107286BL	PBS004	199	-0.0031	-	ISM02.4	PPB	3/13/2018 17:59	SMPL
J1837-01	MF6F08	1534	0.2804	-	ISM02.4	PPB	3/13/2018 18:01	SMPL
J1837-02	MF6F09	1670	0.3092	-	ISM02.4	PPB	3/13/2018 18:03	SMPL
J1837-03	MF6F10	1267	0.2237	-	ISM02.4	PPB	3/13/2018 18:06	SMPL
J1837-04	MF6F12	1671	0.3094	-	ISM02.4	PPB	3/13/2018 18:08	SMPL
J1837-05	MF6F12D	1408	0.2536	-	ISM02.4	PPB	3/13/2018 18:10	SMPL
J1837-06	MF6F12S	13055	2.7267	-	ISM02.4	PPB	3/13/2018 18:13	SMPL
J1837-07	MF6F13	879	0.1413	-	ISM02.4	PPB	3/13/2018 18:15	SMPL
J1837-08	MF6F14	308	0.02	-	ISM02.4	PPB	3/13/2018 18:17	SMPL
J1837-09	MF6F15	314	0.0213	-	ISM02.4	PPB	3/13/2018 18:19	SMPL
J1837-10	MF6F33	649	0.0924	-	ISM02.4	PPB	3/13/2018 18:22	SMPL

LB93879

INSTRUMENT ID : CV1

J1837-11	MF6F34	805	0.1256 -	ISM02.4	PPB	3/13/2018 18:24	SMPL
J1837-12	MF6F35	410	0.0417 -	ISM02.4	PPB	3/13/2018 18:26	SMPL
CCV056	CCV056	23896	5.0286 -	ISM02.4	PPB	3/13/2018 18:29	SMPL
CCB056	CCB056	-67	-0.0596 -	ISM02.4	PPB	3/13/2018 18:31	SMPL
J1837-13	MF6F48	633	0.089 -	ISM02.4	PPB	3/13/2018 18:33	SMPL
J1837-14	MF6F49	321	0.0228 -	ISM02.4	PPB	3/13/2018 18:35	SMPL
J1837-15	MF6F32	666	0.096 -	ISM02.4	PPB	3/13/2018 18:38	SMPL
J1837-18	MF6F36	328	0.0243 -	ISM02.4	PPB	3/13/2018 18:40	SMPL
J1837-19	MF6F47	1582	0.2905 -	ISM02.4	PPB	3/13/2018 18:42	SMPL
CCV057	CCV057	23632	4.9726 -	ISM02.4	PPB	3/13/2018 18:45	SMPL
CCB057	CCB057	-71	-0.0604 -	ISM02.4	PPB	3/13/2018 18:47	SMPL