

Analysis Method: SM4500 S EorF

ANALYST: KETANKUMAR

Parameter: Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB93880

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2276
IODINE SOLUTION .025N 1L	W2283
Starch Solution, 4L	W2121

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB107316BL		1	50	50	5.00	0.00	5.00	5.00	0.00	0.00	0.00	03/13/2018	16:32
2	PB107316BS	25	1	50	50	5.00	0.00	1.66	1.66	3.34	3.34	26.72	03/13/2018	16:35
3	J1900-32		1	50	50	5.00	0.00	3.98	3.98	1.02	1.02	8.16	03/13/2018	16:38

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume