


ISM02.4

LB93886
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42674



A= 0.0000e+000

B= 2.3530e-004

C= -2.5793e-002

Rho= 0.9999864

Accept=Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.021	-0.021	22	0.000	22					-
0.2	0.200	0.222	0.022	1051	0.0 %	1051					11
2.5	2.500	2.520	0.020	10818	0.0 %	10818					1
5.0	5.000	4.976	-0.024	21257	0.0 %	21257					0
7.5	7.500	7.488	-0.012	31934	0.0 %	31934					0
10.0	10.000	10.015	0.015	42674	0.0 %	42674					0

m3
3/13/18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	22	-		0 ISM02.4	PPB	3/13/2018 19:52	Std
	0.2 S0.2	1051	-		0.2 ISM02.4	PPB	3/13/2018 19:54	Std
	2.5 S2.5	10818	-		2.5 ISM02.4	PPB	3/13/2018 19:56	Std
	5 S5	21257	-		5 ISM02.4	PPB	3/13/2018 19:58	Std
	7.5 S7.5	31934	-		7.5 ISM02.4	PPB	3/13/2018 20:03	Std
	10 S10	42674	-		10 ISM02.4	PPB	3/13/2018 20:05	Std
ICV	ICV	16174	3.7799 -		ISM02.4	PPB	3/13/2018 20:08	SMPL
ICB	ICB	-122	-0.0545 -		ISM02.4	PPB	3/13/2018 20:10	SMPL
CCV058	CCV058	21320	4.9908 -		ISM02.4	PPB	3/13/2018 20:12	SMPL
CCB058	CCB058	-122	-0.0545 -		ISM02.4	PPB	3/13/2018 20:14	SMPL
PB107284BL	PBW003	26	-0.0197 -		ISM02.4	PPB	3/13/2018 20:17	SMPL
J1837-16	MF6F51	-100	-0.0493 -		ISM02.4	PPB	3/13/2018 20:19	SMPL
J1837-17	MF6F52	-44	-0.0361 -		ISM02.4	PPB	3/13/2018 20:21	SMPL
CCV059	CCV059	21489	5.0305 -		ISM02.4	PPB	3/13/2018 20:24	SMPL
CCB059	CCB059	-113	-0.0524 -		ISM02.4	PPB	3/13/2018 20:26	SMPL