



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB93890

Temp.: 16.8 °C

SLOPE.: 97.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/13/2018	17:00
2	CAL2	1	Water			7.01	03/13/2018	17:01
3	CAL3	1	Water			10.02	03/13/2018	17:02
4	ICV	1	Water			7.00	03/13/2018	17:05
5	CCV1	1	Water			2.01	03/13/2018	17:06
6	J1910-01	1	Water			6.84	03/13/2018	17:08
7	J1910-01DUP	1	Water			6.85	03/13/2018	17:09
8	J1911-11	1	Water			7.60	03/13/2018	17:10
9	CCV2	1	Water			12.02	03/13/2018	17:11

LB93890

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1910

WorkList ID : 109559

Date : 3/13/2018 4:39:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/13/2018	Water- P	J1910-01	pH	Cool 4 deg C	PSEG01	H42	291	03/13/2018	9040C
03/13/2018	Water- P	J1911-11	pH	Cool 4 deg C	PSEG01	H43	TP-5	03/13/2018	9040C

Date/Time 03-13-18 6:10 (NW)
Received by: JA
Relinquished by: CS

Date/Time 03-13-18 6:10 (NW)
Received by: JA
Relinquished by: JP