



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: Hiral

Run Number: LB93895

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/14/2018	08:40
2	J1881-25		1	>100.0	>212.0	03/14/2018	09:02
3	J1881-25DUP		1	>100.0	>212.0	03/14/2018	09:25
4	J1881-26		1	>100.0	>212.0	03/14/2018	09:45
5	J1881-27		1	>100.0	>212.0	03/14/2018	10:10
6	J1881-28		1	>100.0	>212.0	03/14/2018	10:35
7	J1881-29		1	>100.0	>212.0	03/14/2018	10:50
8	J1881-30		1	>100.0	>212.0	03/14/2018	11:15
9	J1881-31		1	>100.0	>212.0	03/14/2018	11:45
10	J1881-32		1	>100.0	>212.0	03/14/2018	12:05
11	J1881-34		1	>100.0	>212.0	03/14/2018	12:20
12	J1881-35		1	>100.0	>212.0	03/14/2018	12:45
13	J1881-36		1	>100.0	>212.0	03/14/2018	12:58
14	J1881-37		1	>100.0	>212.0	03/14/2018	13:15
15	J1881-38		1	>100.0	>212.0	03/14/2018	13:40
16	J1887-03		1	>100.0	>212.0	03/14/2018	13:56
17	J1887-04		1	>100.0	>212.0	03/14/2018	14:15
18	J1887-05		1	>100.0	>212.0	03/14/2018	14:40
19	J1898-16		1	>100.0	>212.0	03/14/2018	14:50

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 93895

WorkList Name : FP-1887

WorkList ID : 109502

Date : 3/12/2018 11:26:27 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2018	Water	J1881-25 B	Flash Point	Cool 4 deg C	PSEG01	I51	22129	03/09/2018	1010A
03/16/2018	Water	J1881-26 B	Flash Point	Cool 4 deg C	PSEG01	I51	22002	03/09/2018	1010A
03/16/2018	Water	J1881-27 B	Flash Point	Cool 4 deg C	PSEG01	I51	22014	03/09/2018	1010A
03/16/2018	Water	J1881-28 B	Flash Point	Cool 4 deg C	PSEG01	I51	22064	03/09/2018	1010A
03/16/2018	Water	J1881-29 B	Flash Point	Cool 4 deg C	PSEG01	I51	22070	03/09/2018	1010A
03/16/2018	Water	J1881-30 B	Flash Point	Cool 4 deg C	PSEG01	I51	22115	03/09/2018	1010A
03/16/2018	Water	J1881-31 B	Flash Point	Cool 4 deg C	PSEG01	I51	22120	03/09/2018	1010A
03/16/2018	Water	J1881-32 B	Flash Point	Cool 4 deg C	PSEG01	I51	22130	03/09/2018	1010A
03/16/2018	Water	J1881-34 B	Flash Point	Cool 4 deg C	PSEG01	I51	22029	03/09/2018	1010A
03/16/2018	Water	J1881-35 B	Flash Point	Cool 4 deg C	PSEG01	I51	22125	03/09/2018	1010A
03/16/2018	Water	J1881-36 B	Flash Point	Cool 4 deg C	PSEG01	I51	22004	03/09/2018	1010A
03/16/2018	Water	J1881-37 B	Flash Point	Cool 4 deg C	PSEG01	I51	22007	03/09/2018	1010A
03/16/2018	Water	J1881-38 B	Flash Point	Cool 4 deg C	PSEG01	I51	22028	03/09/2018	1010A
03/16/2018	Water	J1887-03 B	Flash Point	Cool 4 deg C	PSEG01	I51	16662-17600	03/09/2018	1010A
03/16/2018	Water	J1887-04 B	Flash Point	Cool 4 deg C	PSEG01	I51	16626	03/09/2018	1010A
03/16/2018	Water	J1887-05 B	Flash Point	Cool 4 deg C	PSEG01	I51	13572	03/09/2018	1010A
03/15/2018	Water	J1898-16 C	Flash Point	Cool 4 deg C	PSEG01	I53	TP-11	03/12/2018	1010A

Date/Time 3/14/18 8:15 AM
 Received by: KS
 Relinquished by: CS

Date/Time 3/14/18 14:55
 Received by: CS
 Relinquished by: KS