

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/14/2018 13:20

Test: ^{Reactive} ~~Total~~ CN Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.339	0.0	0.094	
ICB1	0.579	0.0	0.004	
CCV1	241.874	0.0	0.232	
CCB1	0.336	0.0	0.004	
PB107347BL	0.203	0.0	0.003	
J1749-04	0.128	0.0	0.003	
J1749-04DUP	0.234	0.0	0.004	
J1754-02	0.015	0.0	0.003	
J1754-04	0.041	0.0	0.003	
J1754-06	0.123	0.0	0.003	
J1754-08	0.114	0.0	0.003	
J1754-10	0.494	0.0	0.004	
J1908-04	0.378	0.0	0.004	
J1908-08	0.490	0.0	0.004	
CCV2	240.626	0.0	0.231	
CCB2	0.524	0.0	0.004	
J1909-02	0.390	0.0	0.004	
J1911-03	0.383	0.0	0.004	
J1911-06	0.272	0.0	0.004	
J1911-10	0.305	0.0	0.004	
PB107348BL	0.345	0.0	0.004	
J1910-02	0.264	0.0	0.004	
J1910-02DUP	0.243	0.0	0.004	
J1911-12	0.445	0.0	0.004	
CCV3	255.512	0.0	0.245	
CCB3	0.344	0.0	0.004	

N 26
Mean 32.346
SD 80.9371
CV% 250.22

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Mar 14 13:02:18 2018

Wed Mar 14 13:14:13 2018

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.2234	µg/l	3/14/2018 10:59:04	
5.OPPBCN	A	Reactive CN	P	5.0602	µg/l	3/14/2018 10:59:05	
10PPBCN	A	Reactive CN	P	10.1223	µg/l	3/14/2018 10:59:06	
50PPBCN	A	Reactive CN	P	49.4104	µg/l	3/14/2018 10:59:07	
100PPBCN	A	Reactive CN	P	100.4242	µg/l	3/14/2018 10:59:08	
250PPBCN	A	Reactive CN	P	249.577	µg/l	3/14/2018 10:59:09	
500PPBCN	A	Reactive CN	P	500.1826	µg/l	3/14/2018 10:59:10	
ICV1	S	Reactive CN	P	96.3386	µg/l	3/14/2018 13:02:18	
ICB1	S	Reactive CN	P	0.5795	µg/l	3/14/2018 13:02:19	
CCV1	S	Reactive CN	P	241.8739	µg/l	3/14/2018 13:02:20	
CCB1	S	Reactive CN	P	0.3359	µg/l	3/14/2018 13:02:21	
PB107347BL	S	Reactive CN	P	0.2033	µg/l	3/14/2018 13:02:22	
J1749-04	S	Reactive CN	P	0.1276	µg/l	3/14/2018 13:02:23	
J1749-04DUP	S	Reactive CN	P	0.2342	µg/l	3/14/2018 13:02:24	
J1754-02	S	Reactive CN	P	0.0148	µg/l	3/14/2018 13:02:25	
J1754-04	S	Reactive CN	P	0.0413	µg/l	3/14/2018 13:02:26	
J1754-06	S	Reactive CN	P	0.1234	µg/l	3/14/2018 13:02:27	
J1754-08	S	Reactive CN	P	0.114	µg/l	3/14/2018 13:02:28	
J1754-10	S	Reactive CN	P	0.4945	µg/l	3/14/2018 13:09:51	
J1908-04	S	Reactive CN	P	0.3777	µg/l	3/14/2018 13:09:52	
J1908-08	S	Reactive CN	P	0.4901	µg/l	3/14/2018 13:09:53	
CCV2	S	Reactive CN	P	240.6255	µg/l	3/14/2018 13:09:54	
CCB2	S	Reactive CN	P	0.5236	µg/l	3/14/2018 13:09:55	
J1909-02	S	Reactive CN	P	0.3905	µg/l	3/14/2018 13:09:56	
J1911-03	S	Reactive CN	P	0.3832	µg/l	3/14/2018 13:09:57	
J1911-06	S	Reactive CN	P	0.2721	µg/l	3/14/2018 13:09:58	
J1911-10	S	Reactive CN	P	0.3046	µg/l	3/14/2018 13:09:59	
PB107348BL	S	Reactive CN	P	0.3452	µg/l	3/14/2018 13:10:00	
J1910-02	S	Reactive CN	P	0.2641	µg/l	3/14/2018 13:10:01	
J1910-02DUP	S	Reactive CN	P	0.2427	µg/l	3/14/2018 13:14:10	
J1911-12	S	Reactive CN	P	0.4448	µg/l	3/14/2018 13:14:11	
CCV3	S	Reactive CN	P	255.5121	µg/l	3/14/2018 13:14:12	
CCB3	S	Reactive CN	P	0.3436	µg/l	3/14/2018 13:14:13	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/14/2018 10:59


Test Total CN

Accepted 3/14/2018 10:59

Factor 1059
Bias 0.003

Coeff. of det. 0.999996

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.2234	0.0000	
2	5.0PPBCN	0.008	5.0602	5.0000	
3	10PPBCN	0.013	10.1223	10.0000	
4	50PPBCN	0.050	49.4104	50.0000	
5	100PPBCN	0.098	100.4242	100.0000	
6	250PPBCN	0.239	249.5770	250.0000	
7	500PPBCN	0.476	500.1826	500.0000	