


7471B

LB93909
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45002



A= 0.0000e+000

B= 2.2512e-004

C= 1.3363e-002

Rho= 0.9996361

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.036	0.036	102	0.000	102				
0.2	0.200	0.220	0.020	917	0.0 %	917				
2.5	2.500	2.511	0.011	11095	0.0 %	11095				
5.0	5.000	4.979	-0.021	22058	0.0 %	22058				
7.5	7.500	7.309	-0.191	32409	0.0 %	32409				
10	10.000	10.144	0.144	45002	0.0 %	45002				

LB93909 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Mar 2018 15:10:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	102	-	0.0000	7471B	PPB	14 Mar 2018 15:13:04	S	Std
0.2 - 1	50.2	917	-	0.2000	7471B	PPB	14 Mar 2018 15:15:22	S	Std
2.5 - 1	52.5	11095	-	2.5000	7471B	PPB	14 Mar 2018 15:17:40	S	Std
5.0 - 1	55	22058	-	5.0000	7471B	PPB	14 Mar 2018 15:19:57	S	Std
7.5 - 1	57.5	32409	-	7.5000	7471B	PPB	14 Mar 2018 15:22:14	S	Std
10 - 1	510	45002	-	10.0000	7471B	PPB	14 Mar 2018 15:24:31	S	Std
ICV43 - 1	ICV43	17271	3.9015	-7471B	PPB	PPB	14 Mar 2018 15:30:14	U	SMPL
ICB43 - 1	ICB43	-21	0.0086	-7471B	PPB	PPB	14 Mar 2018 15:32:30	U	SMPL
CCV67 - 1	CCV67	22566	5.0935	-7471B	PPB	PPB	14 Mar 2018 15:34:45	U	SMPL
CCB67 - 1	CCB67	-32	0.0062	-7471B	PPB	PPB	14 Mar 2018 15:37:01	U	SMPL
CRA - 1	CRA	867	0.2085	-7471B	PPB	PPB	14 Mar 2018 15:39:18	U	SMPL
HighStd - 1	HighStd	45192	10.1871	-7471B	PPB	PPB	14 Mar 2018 15:41:34	U	SMPL
ChkStd - 1	ChkStd	29948	6.7553	-7471B	PPB	PPB	14 Mar 2018 15:43:50	U	SMPL
PB107287BL - 1	PBS	-198	-0.0312	-7471B	PPB	PPB	14 Mar 2018 15:46:09	U	SMPL
PB107287BS - 1	LCSS	15251	3.4467	-7471B	PPB	PPB	14 Mar 2018 15:48:27	U	SMPL
J1749-03 - 1	BU-02-031318	2274	0.5253	-7471B	PPB	PPB	14 Mar 2018 15:50:44	U	SMPL
J1753-01 - 1	HD-02-031218	479	0.1212	-7471B	PPB	PPB	14 Mar 2018 15:53:01	U	SMPL
J1754-01 - 1	CL-01-031318-A	3015	0.6921	-7471B	PPB	PPB	14 Mar 2018 15:55:18	U	SMPL
J1754-03 - 1	CL-01-031318-B	2334	0.5388	-7471B	PPB	PPB	14 Mar 2018 15:57:36	U	SMPL
J1754-05 - 1	CL-01-031318-C	3150	0.7225	-7471B	PPB	PPB	14 Mar 2018 15:59:52	U	SMPL
CCV68 - 1	CCV68	21951	4.9550	-7471B	PPB	PPB	14 Mar 2018 16:02:08	U	SMPL
CCB68 - 1	CCB68	9	0.0154	-7471B	PPB	PPB	14 Mar 2018 16:04:24	U	SMPL
J1754-07 - 1	CL-01-031318-D	905	0.2171	-7471B	PPB	PPB	14 Mar 2018 16:06:41	U	SMPL
J1754-09 - 1	CL-01-031318-E	2083	0.4823	-7471B	PPB	PPB	14 Mar 2018 16:08:57	U	SMPL
J1890-01 - 1	DRUM	40712	9.1785	-7471B	PPB	PPB	14 Mar 2018 16:11:13	U	SMPL
J1893-01 - 1	TP-2	4759	1.0847	-7471B	PPB	PPB	14 Mar 2018 16:13:29	U	SMPL
J1898-03 - 1	TP-10-COMP	2054	0.4758	-7471B	PPB	PPB	14 Mar 2018 16:15:48	U	SMPL
J1898-07 - 1	TP-11A-COMP	304	0.0818	-7471B	PPB	PPB	14 Mar 2018 16:18:05	U	SMPL
J1898-09 - 1	TP-11B-COMP	457	0.1162	-7471B	PPB	PPB	14 Mar 2018 16:20:22	U	SMPL
J1898-13 - 1	TP-12-COMP	799	0.1932	-7471B	PPB	PPB	14 Mar 2018 16:22:40	U	SMPL
J1898-13DUP - 1	TP-12-COMPDUP	842	0.2029	-7471B	PPB	PPB	14 Mar 2018 16:24:58	U	SMPL
J1898-13MS - 1	TP-12-COMPMS	16988	3.8377	-7471B	PPB	PPB	14 Mar 2018 16:27:15	U	SMPL
CCV69 - 1	CCV69	22430	5.0629	-7471B	PPB	PPB	14 Mar 2018 16:29:31	U	SMPL
CCB69 - 1	CCB69	-40	0.0044	-7471B	PPB	PPB	14 Mar 2018 16:31:47	U	SMPL
J1898-13MSD - 1	TP-12-COMPMSD	16959	3.8312	-7471B	PPB	PPB	14 Mar 2018 16:34:03	U	SMPL
J1899-01 - 1	EP-32A	550	0.1372	-7471B	PPB	PPB	14 Mar 2018 16:36:19	U	SMPL
J1908-01 - 1	TP-A	567	0.1410	-7471B	PPB	PPB	14 Mar 2018 16:38:35	U	SMPL
J1908-05 - 1	TP-B	2052	0.4753	-7471B	PPB	PPB	14 Mar 2018 16:40:51	U	SMPL
J1909-01 - 1	1-3-RO-COMP	1493	0.3495	-7471B	PPB	PPB	14 Mar 2018 16:43:08	U	SMPL
J1911-02 - 1	TP-5A-COMP	361	0.0946	-7471B	PPB	PPB	14 Mar 2018 16:45:24	U	SMPL
J1911-05 - 1	TP-5B-COMP	328	0.0872	-7471B	PPB	PPB	14 Mar 2018 16:47:41	U	SMPL
J1911-09 - 1	TP-13-COMP	1585	0.3702	-7471B	PPB	PPB	14 Mar 2018 16:49:59	U	SMPL
J1898-13LX5 - 1		184	0.0548	-7471B	PPB	PPB	14 Mar 2018 16:52:17	U	SMPL
J1898-13A - 1		17912	4.0458	-7471B	PPB	PPB	14 Mar 2018 16:54:33	U	SMPL
CCV70 - 1	CCV70	22493	5.0770	-7471B	PPB	PPB	14 Mar 2018 16:56:49	U	SMPL
CCB70 - 1	CCB70	-49	0.0023	-7471B	PPB	PPB	14 Mar 2018 16:59:06	U	SMPL