



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93912

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/14/2018	17:35
2	CAL2	1	Water			7.00	03/14/2018	17:36
3	CAL3	1	Water			10.03	03/14/2018	17:37
4	ICV	1	Water			7.01	03/14/2018	17:39
5	CCV1	1	Water			2.01	03/14/2018	17:40
6	J1752-02	1	Solid	20.02	20	6.61	03/14/2018	17:41
7	J1757-02	1	Solid	20.04	20	8.92	03/14/2018	17:42
8	J1757-04	1	Solid	20.02	20	8.68	03/14/2018	17:43
9	J1757-06	1	Solid	20.04	20	8.45	03/14/2018	17:45
10	J1929-04	1	Solid	20.02	20	7.32	03/14/2018	17:46
11	J1929-08	1	Solid	20.01	20	8.06	03/14/2018	17:48
12	J1929-12	1	Solid	20.04	20	6.62	03/14/2018	17:49
13	J1929-12DUP	1	Solid	20.03	20	6.63	03/14/2018	17:50
14	CCV2	1	Water			12.02	03/14/2018	17:51

1893912

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J1929 WorkList ID : 109613 Date : 3/14/2018 5:13:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/14/2018	Solid	J1752-02	Corrosivity	Cool 4 deg C	PSEG05	I62	NB-08-031418	03/14/2018	9045C
03/14/2018	Solid	J1757-02	Corrosivity	Cool 4 deg C	PSEG05	I11	EO-01-031418-A	03/14/2018	9045C
03/14/2018	Solid	J1757-04	Corrosivity	Cool 4 deg C	PSEG05	I11	EO-01-031418-B	03/14/2018	9045C
03/14/2018	Solid	J1757-06	Corrosivity	Cool 4 deg C	PSEG05	I11	EO-01-031418-C	03/14/2018	9045C
03/14/2018	Solid	J1929-04	Corrosivity	Cool 4 deg C	PSEG01	I63	LS-PIT-1	03/14/2018	9045C
03/14/2018	Solid	J1929-08	Corrosivity	Cool 4 deg C	PSEG01	I63	LS-PIT-2	03/14/2018	9045C
03/14/2018	Solid	J1929-12	Corrosivity	Cool 4 deg C	PSEG01	I63	LS-PIT-3	03/14/2018	9045C

Date/Time 03/14/18 5:13 PM
 Received by: JD
 Relinquished by: JD

Date/Time 03/14/18 6:40 PM
 Received by: JD
 Relinquished by: JD