


7471B

LB93921
INSTRUMENT ID: CV2

Linear

 μ Abs.:

43904



A= 0.0000e+000

B= 2.2632e-004

C= 2.8688e-002

Rho= 0.9995671

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	-45	0.000	-45				
0.2	0.200	0.257	0.057	1007	0.0 %	1007				
2.5	2.500	2.482	-0.018	10842	0.0 %	10842				
5.0	5.000	4.808	-0.192	21116	0.0 %	21116				
7.5	7.500	7.669	0.169	33760	0.0 %	33760				
10	10.000	9.965	-0.035	43904	0.0 %	43904				

LB93921 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Mar 2018 09:50:27

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-45	-	0.00007471B	PPB	15 Mar 2018 09:55:26	S	Std
0.2 - 1	50.2	1007	-	0.20007471B	PPB	15 Mar 2018 09:57:43	S	Std
2.5 - 1	52.5	10842	-	2.50007471B	PPB	15 Mar 2018 10:00:01	S	Std
5.0 - 1	55	21116	-	5.00007471B	PPB	15 Mar 2018 10:02:18	S	Std
7.5 - 1	57.5	33760	-	7.50007471B	PPB	15 Mar 2018 10:04:35	S	Std
10 - 1	510	43904	-	10.00007471B	PPB	15 Mar 2018 10:10:12	S	Std
ICV45 - 1	ICV45	17159	3.9122	-7471B	PPB	15 Mar 2018 10:13:28	U	SMPL
ICB45 - 1	ICB45	-182	-0.0125	-7471B	PPB	15 Mar 2018 10:15:43	U	SMPL
CCV74 - 1	CCV74	21664	4.9318	-7471B	PPB	15 Mar 2018 10:17:59	U	SMPL
CCB74 - 1	CCB74	-235	-0.0245	-7471B	PPB	15 Mar 2018 10:20:15	U	SMPL
CRA - 1	CRA	977	0.2498	-7471B	PPB	15 Mar 2018 10:22:32	U	SMPL
HighStd - 1	HighStd	44594	10.1214	-7471B	PPB	15 Mar 2018 10:24:47	U	SMPL
ChkStd - 1	ChkStd	30568	6.9470	-7471B	PPB	15 Mar 2018 10:27:04	U	SMPL
PB107375BL - 1	PBS	-246	-0.0270	-7471B	PPB	15 Mar 2018 10:29:22	U	SMPL
PB107375BS - 1	LCSS	15299	3.4912	-7471B	PPB	15 Mar 2018 10:31:40	U	SMPL
J1752-01 - 1	NB-08-031418	2086	0.5008	-7471B	PPB	15 Mar 2018 10:33:57	U	SMPL
J1757-01 - 1	EO-01-031418-A	2554	0.6067	-7471B	PPB	15 Mar 2018 10:36:14	U	SMPL
J1757-03 - 1	EO-01-031418-B	2444	0.5818	-7471B	PPB	15 Mar 2018 10:38:32	U	SMPL
J1757-05 - 1	EO-01-031418-C	1779	0.4313	-7471B	PPB	15 Mar 2018 10:40:49	U	SMPL
J1915-01 - 1	P002-S001-0001-01	2403	0.5725	-7471B	PPB	15 Mar 2018 10:43:06	U	SMPL
CCV75 - 1	CCV75	21488	4.8919	-7471B	PPB	15 Mar 2018 10:45:22	U	SMPL
CCB75 - 1	CCB75	-176	-0.0111	-7471B	PPB	15 Mar 2018 10:47:38	U	SMPL
J1915-01DUP - 1	P002-S001-0001-01DUP	2639	0.6260	-7471B	PPB	15 Mar 2018 10:49:54	U	SMPL
J1915-02 - 1	J1915-01MS	17758	4.0478	-7471B	PPB	15 Mar 2018 10:52:10	U	SMPL
J1915-03 - 1	J1915-01MSD	17871	4.0733	-7471B	PPB	15 Mar 2018 10:54:27	U	SMPL
J1915-04 - 1	P002-S001-0001-02	2330	0.5560	-7471B	PPB	15 Mar 2018 10:56:43	U	SMPL
J1927-01 - 1	FK-BPM-01-C01-C01A-D01-D01A	1285	0.3195	-7471B	PPB	15 Mar 2018 10:59:00	U	SMPL
J1928-01 - 1	FK-BPM-01-E01-E01A-F01-F01A	1238	0.3089	-7471B	PPB	15 Mar 2018 11:01:17	U	SMPL
J1929-01 - 1	LS-PIT-1	5906	1.3654	-7471B	PPB	15 Mar 2018 11:03:34	U	SMPL
J1929-05 - 1	LS-PIT-2	15076	3.4408	-7471B	PPB	15 Mar 2018 11:05:52	U	SMPL
J1929-09 - 1	LS-PIT-3	2065	0.4960	-7471B	PPB	15 Mar 2018 11:08:10	U	SMPL
J1930-01 - 1	VALLEY-RD	5118	1.1870	-7471B	PPB	15 Mar 2018 11:10:26	U	SMPL
CCV76 - 1	CCV76	21466	4.8870	-7471B	PPB	15 Mar 2018 11:12:42	U	SMPL
CCB76 - 1	CCB76	-176	-0.0111	-7471B	PPB	15 Mar 2018 11:14:59	U	SMPL
J1915-01LX5 - 1		381	0.1149	-7471B	PPB	15 Mar 2018 11:17:15	U	SMPL
J1915-01A - 1		19107	4.3531	-7471B	PPB	15 Mar 2018 11:19:32	U	SMPL
CCV77 - 1	CCV77	21618	4.9214	-7471B	PPB	15 Mar 2018 11:21:48	U	SMPL
CCB77 - 1	CCB77	-178	-0.0116	-7471B	PPB	15 Mar 2018 11:24:04	U	SMPL