



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93925

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP62745
Residual chlorine std, 0PPM	WP62736
Residual chlorine std, 0.1PPM	WP62740
Residual chlorine std, 0.2PPM	WP62741
Residual chlorine std, and CCV , 0.4PPM	WP62742
Residual chlorine std, 1.6PPM	WP62743
Residual chlorine ICV-LCS, 0.4PPM	WP62744
Residual chlorine LOD, 0.05PPM	WP62738
Total Chlorine Powder Pillows	W1960

Intercept: 0.0006

Slope: 1.0699

Regression: 0.999877

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	03/15/2018	11:00
2	CAL2	0.1	1	0.000	0.120	0.120	03/15/2018	11:01
3	CAL3	0.2	1	0.000	0.210	0.210	03/15/2018	11:01
4	CAL4	0.4	1	0.000	0.430	0.430	03/15/2018	11:02
5	CAL5	0.8	1	0.000	0.840	0.840	03/15/2018	11:03
6	CAL6	1.6	1	0.000	1.720	1.720	03/15/2018	11:03



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB93925

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.440	0.440	0.411	03/15/2018	11:04
2	ICB				1	0.000	0.000	0.000	-0.001	03/15/2018	11:04
3	CCV1			0.4	1	0.000	0.440	0.440	0.411	03/15/2018	11:05
4	CCB1				1	0.000	0.000	0.000	-0.001	03/15/2018	11:05
5	LB93925BL	50	50		1	0.000	0.000	0.000	-0.001	03/15/2018	11:06
6	LB93925BS	50	50	0.4	1	0.000	0.440	0.440	0.411	03/15/2018	11:06
7	J1161-07	50	50		1	0.000	0.060	0.060	0.056	03/15/2018	11:07
8	J1161-09	50	50		1	0.000	0.070	0.070	0.065	03/15/2018	11:07
9	J1900-19	50	50		1	0.000	2.500	2.500	2.336	03/15/2018	11:08
10	J1900-19				5	0.000	0.540	0.540	0.504	03/15/2018	11:15
11	CCV2			0.4	1	0.000	0.410	0.410	0.383	03/15/2018	11:15
12	CCB2				1	0.000	0.000	0.000	-0.001	03/15/2018	11:16