

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/15/2018 11:41

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.964	0.0	0.113	
ICB1	0.005	0.0	0.011	
CCV1	1.041	0.0	0.121	
CCB1	0.011	0.0	0.011	
PB107362BL	0.009	0.0	0.011	
PB107362BS	0.964	0.0	0.112	
J1900-26	2.483	0.0	0.274	Test limit high
CCV2	1.047	0.0	0.121	
CCB2	0.010	0.0	0.011	
J1900-26DLX2	1.240	0.0	0.142	
CCV3	1.021	0.0	0.119	
CCB3	-0.002	0.0	0.010	

N 12
Mean 0.733
SD 0.7592
CV% 103.61

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Mar 15 09:48:45 2018

Thu Mar 15 11:39:52 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0522	mg/l	3/15/2018 10:11:44	
0.1PPM	A	Phenolics-	P	0.1012	mg/l	3/15/2018 10:11:45	
0.25PPM	A	Phenolics-	P	0.2497	mg/l	3/15/2018 10:11:46	
0.50PPM	A	Phenolics-	P	0.499	mg/l	3/15/2018 10:11:47	
1.0PPM	A	Phenolics-	P	0.9955	mg/l	3/15/2018 10:11:48	
2.0PPM	A	Phenolics-	P	2.0024	mg/l	3/15/2018 10:11:49	
ICV1	S	Phenolics-	P	0.9643	mg/l	3/15/2018 11:20:07	
ICB1	S	Phenolics-	P	0.0047	mg/l	3/15/2018 11:20:08	
CCV1	S	Phenolics-	P	1.0407	mg/l	3/15/2018 11:20:09	
CCB1	S	Phenolics-	P	0.0111	mg/l	3/15/2018 11:20:10	
PB107362BL	S	Phenolics-	P	0.0093	mg/l	3/15/2018 11:20:11	
PB107362BS	S	Phenolics-	P	0.9635	mg/l	3/15/2018 11:20:12	
J1900-26	S	Phenolics-	P	2.4831	mg/l	3/15/2018 11:20:13	
CCV2	S	Phenolics-	P	1.047	mg/l	3/15/2018 11:20:16	
CCB2	S	Phenolics-	P	0.01	mg/l	3/15/2018 11:20:17	
J1900-26DLX2	S	Phenolics-	P	1.2403	mg/l	3/15/2018 11:39:50	
CCV3	S	Phenolics-	P	1.0211	mg/l	3/15/2018 11:39:51	
CCB3	S	Phenolics-	P	-0.0017	mg/l	3/15/2018 11:39:52	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/15/2018 10:12


Test Phenolics-

Accepted 3/15/2018 10:12

Factor 9.423
Bias 0.010

Coeff. of det. 0.999988

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.016	0.0522	0.0500	
2	Phenol-2	0.021	0.1012	0.1000	
3	Phenol-2	0.037	0.2497	0.2500	
4	Phenol-2	0.063	0.4990	0.5000	
5	Phenol-2	0.116	0.9955	1.0000	
6	Phenol-2	0.223	2.0024	2.0000	