


7470A

LB 93932
INSTRUMENT ID: C42

Linear

 μ Abs.:

35476



A= 0.0000e+000

B= 2.8392e-004

C= 9.6977e-003

Rho= 0.9998433

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	-66	0.000	-66				
0.2	0.200	0.210	0.010	707	0.0 %	707				
2.5	2.500	2.543	0.043	8923	0.0 %	8923				
5.0	5.000	5.003	0.003	17586	0.0 %	17586				
7.5	7.500	7.370	-0.130	25925	0.0 %	25925				
10.0	10.000	10.082	0.082	35476	0.0 %	35476				

LB93932 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Mar 2018 14:26:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-66	-	0.0000	7470A	PPB	15 Mar 2018 14:28:44	S	Std
0.2 - 1	50.2	707	-	0.2000	7470A	PPB	15 Mar 2018 14:30:53	S	Std
2.5 - 1	52.5	8923	-	2.5000	7470A	PPB	15 Mar 2018 14:33:01	S	Std
5.0 - 1	55	17586	-	5.0000	7470A	PPB	15 Mar 2018 14:35:09	S	Std
7.5 - 1	57.5	25925	-	7.5000	7470A	PPB	15 Mar 2018 14:37:17	S	Std
10.0 - 1	510	35476	-	10.0000	7470A	PPB	15 Mar 2018 14:41:58	S	Std
ICV48 - 1	ICV48	14374	4.0908	-	7470A	PPB	15 Mar 2018 14:44:35	U	SMPL
ICB48 - 1	ICB48	-249	-0.0610	-	7470A	PPB	15 Mar 2018 14:46:41	U	SMPL
CCV78 - 1	CCV78	17725	5.0423	-	7470A	PPB	15 Mar 2018 14:48:48	U	SMPL
CCB78 - 1	CCB78	-239	-0.0582	-	7470A	PPB	15 Mar 2018 14:50:55	U	SMPL
CRA - 1	CRA	693	0.2065	-	7470A	PPB	15 Mar 2018 14:53:02	U	SMPL
HighStd - 1	HighStd	35539	10.1001	-	7470A	PPB	15 Mar 2018 14:55:09	U	SMPL
ChkStd - 1	ChkStd	24942	7.0913	-	7470A	PPB	15 Mar 2018 14:57:16	U	SMPL
PB107363BL - 1	PBW	-222	-0.0533	-	7470A	PPB	15 Mar 2018 14:59:25	U	SMPL
PB107363BS - 1	LCSW	13926	3.9636	-	7470A	PPB	15 Mar 2018 15:01:34	U	SMPL
PB107363TB - 1	PB107318TB	-98	-0.0181	-	7470A	PPB	15 Mar 2018 15:03:42	U	SMPL
J1749-04 - 1	BU-02-031318	-20	0.0040	-	7470A	PPB	15 Mar 2018 15:05:50	U	SMPL
J1754-02 - 1	CL-01-031318-A	17	0.0145	-	7470A	PPB	15 Mar 2018 15:07:58	U	SMPL
J1754-04 - 1	CL-01-031318-B	1	0.0100	-	7470A	PPB	15 Mar 2018 15:10:07	U	SMPL
J1754-06 - 1	CL-01-031318-C	7	0.0117	-	7470A	PPB	15 Mar 2018 15:12:14	U	SMPL
CCV79 - 1	CCV79	17775	5.0565	-	7470A	PPB	15 Mar 2018 15:14:21	U	SMPL
CCB79 - 1	CCB79	-242	-0.0590	-	7470A	PPB	15 Mar 2018 15:16:27	U	SMPL
J1754-08 - 1	CL-01-031318-D	-1	0.0094	-	7470A	PPB	15 Mar 2018 15:18:36	U	SMPL
J1754-10 - 1	CL-01-031318-E	94	0.0364	-	7470A	PPB	15 Mar 2018 15:25:30	U	SMPL
J1754-10DUP - 1	CL-01-031318-EDUP	106	0.0398	-	7470A	PPB	15 Mar 2018 15:27:37	U	SMPL
J1754-10MS - 1	CL-01-031318-EMS	13507	3.8447	-	7470A	PPB	15 Mar 2018 15:29:44	U	SMPL
J1754-10MSD - 1	CL-01-031318-EMSD	14371	4.0900	-	7470A	PPB	15 Mar 2018 15:31:51	U	SMPL
J1908-04 - 1	TP-A	-109	-0.0213	-	7470A	PPB	15 Mar 2018 15:33:59	U	SMPL
J1908-08 - 1	TP-B	-23	0.0032	-	7470A	PPB	15 Mar 2018 15:36:07	U	SMPL
J1909-02 - 1	1-3-RO-COMP	0	0.0097	-	7470A	PPB	15 Mar 2018 15:38:16	U	SMPL
J1911-03 - 1	TP-5A-COMP	8	0.0120	-	7470A	PPB	15 Mar 2018 15:40:25	U	SMPL
J1911-06 - 1	TP-5B-COMP	32	0.0188	-	7470A	PPB	15 Mar 2018 15:42:32	U	SMPL
CCV80 - 1	CCV80	17795	5.0621	-	7470A	PPB	15 Mar 2018 15:44:38	U	SMPL
CCB80 - 1	CCB80	-219	-0.0525	-	7470A	PPB	15 Mar 2018 15:46:45	U	SMPL
J1911-10 - 1	TP-13-COMP	-3	0.0088	-	7470A	PPB	15 Mar 2018 15:48:53	U	SMPL
PB107364BL - 1	PBW	-12	0.0063	-	7470A	PPB	15 Mar 2018 15:51:00	U	SMPL
PB107364BS - 1	LCSW	14274	4.0624	-	7470A	PPB	15 Mar 2018 15:53:07	U	SMPL
PB107364TB - 1	PB107324TB	-141	-0.0303	-	7470A	PPB	15 Mar 2018 15:55:14	U	SMPL
J1910-02 - 1	291	-30	0.0012	-	7470A	PPB	15 Mar 2018 15:57:21	U	SMPL
J1910-02DUP - 1	291DUP	-9	0.0071	-	7470A	PPB	15 Mar 2018 15:59:28	U	SMPL
J1910-02MS - 1	291MS	15076	4.2901	-	7470A	PPB	15 Mar 2018 16:01:37	U	SMPL
J1910-02MSD - 1	291MSD	14840	4.2231	-	7470A	PPB	15 Mar 2018 16:03:45	U	SMPL
J1911-12 - 1	TP-5	-133	-0.0281	-	7470A	PPB	15 Mar 2018 16:05:54	U	SMPL
CCV81 - 1	CCV81	18017	5.1252	-	7470A	PPB	15 Mar 2018 16:08:01	U	SMPL
CCB81 - 1	CCB81	-232	-0.0562	-	7470A	PPB	15 Mar 2018 16:10:09	U	SMPL
J1754-10LX5 - 1		-67	-0.0093	-	7470A	PPB	15 Mar 2018 16:12:17	U	SMPL
J1754-10A - 1		13595	3.8697	-	7470A	PPB	15 Mar 2018 16:14:24	U	SMPL
J1910-02LX5 - 1		-191	-0.0445	-	7470A	PPB	15 Mar 2018 16:16:31	U	SMPL
J1910-02A - 1		13564	3.8609	-	7470A	PPB	15 Mar 2018 16:18:38	U	SMPL
CCV82 - 1	CCV82	17694	5.0335	-	7470A	PPB	15 Mar 2018 16:20:46	U	SMPL
CCB82 - 1	CCB82	-237	-0.0576	-	7470A	PPB	15 Mar 2018 16:22:53	U	SMPL