



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93940

Temp.: 22.2 °C

BalanceID: WC SC-4

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/15/2018	17:15
2	CAL2	1	Water			7.00	03/15/2018	17:16
3	CAL3	1	Water			10.03	03/15/2018	17:18
4	ICV	1	Water			7.01	03/15/2018	17:19
5	CCV1	1	Water			2.01	03/15/2018	17:20
6	J1750-02	1	Solid	20.02	20	6.94	03/15/2018	17:21
7	J1750-02DUP	1	Solid	20	20	6.95	03/15/2018	17:22
8	J1750-04	1	Solid	20.03	20	7.32	03/15/2018	17:24
9	J1750-06	1	Solid	20.05	20	8.17	03/15/2018	17:25
10	J1756-06	1	Solid	20.01	20	7.29	03/15/2018	17:26
11	J1938-04	1	Solid	20.03	20	8.12	03/15/2018	17:27
12	J1938-08	1	Solid	20.05	20	8.04	03/15/2018	17:30
13	J1938-12	1	Solid	20.02	20	8.20	03/15/2018	17:31
14	J1940-02	1	Solid	20.03	20	8.40	03/15/2018	17:32
15	J1945-04	1	Solid	20.02	20	8.64	03/15/2018	17:33
16	CCV2	1	Water			12.02	03/15/2018	17:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1938

WorkList ID : 109652

Date : 3/15/2018 2:16:05 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/15/2018	Solid	J1750-02	Corrosivity	Cool 4 deg C	PSEG05	I63	HR-05-031518-A	03/15/2018	9045C
03/15/2018	Solid	J1750-04	Corrosivity	Cool 4 deg C	PSEG05	I63	HR-05-031518-B	03/15/2018	9045C
03/15/2018	Solid	J1750-06	Corrosivity	Cool 4 deg C	PSEG05	I63	HR-05-031518-C	03/15/2018	9045C
03/15/2018	Solid	J1756-06	Corrosivity	Cool 4 deg C	PSEG05		OR-02-031518	03/15/2018	9045C
03/15/2018	Solid	J1938-04	Corrosivity	Cool 4 deg C	PSEG01	I51	CS-1	03/15/2018	9045C
03/15/2018	Solid	J1938-08	Corrosivity	Cool 4 deg C	PSEG01	I51	CS-2	03/15/2018	9045C
03/15/2018	Solid	J1938-12	Corrosivity	Cool 4 deg C	PSEG01	I51	CS-3	03/15/2018	9045C
03/15/2018	Solid	J1940-02	Corrosivity	Cool 4 deg C	PSEG01	I21	SP-1-5	03/15/2018	9045C
03/15/2018	Solid	J1945-04	Corrosivity	Cool 4 deg C	PSEG01	I62	SP-A	03/15/2018	9045C

Date/Time 03-15-18 2:30 PM
 Received by: JD
 Relinquished by: [Signature]

Date/Time 03-15-18 6:00 PM
 Received by: [Signature]
 Relinquished by: [Signature]