



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB93950

Temp.: 16.6 °C

SLOPE.: 97.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/16/2018	12:10
2	CAL2	1	Water			7.00	03/16/2018	12:11
3	CAL3	1	Water			10.02	03/16/2018	12:12
4	ICV	1	Water			7.01	03/16/2018	12:14
5	CCV1	1	Water			2.01	03/16/2018	12:15
6	J1931-18	1	Water			11.56	03/16/2018	12:16
7	J1931-18DUP	1	Water			11.57	03/15/2018	12:18
8	J1931-19	1	Water			4.52	03/16/2018	12:19
9	CCV2	1	Water			12.01	03/16/2018	12:20

2393950

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph j1931

WorkList ID : 109685

Date : 3/16/2018 11:59:49 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/14/2018	Water	J1931-18-13	pH	Cool 4 deg C	PSEG02	I61	1472-75-WW	03/14/2018	9045C
03/14/2018	Water	J1931-19-13	pH	Cool 4 deg C	PSEG02	I61	1482-1484-87-OILY-SNOW	03/14/2018	9045C

Date/Time 03-16-18 12:00 PM
Received by: JH
Relinquished by: SP

Date/Time 03-16-18 12:35 PM
Received by: SP
Relinquished by: SP