


7471B

LB 93958
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41485



A= 0.0000e+000

B= 2.4220e-004

C= -3.1194e-002

Rho= 0.9998800

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	46	0.000	46				
0.2	0.200	0.203	0.003	966	0.0 %	966				
2.5	2.500	2.486	-0.014	10394	0.0 %	10394				
5.0	5.000	5.104	0.104	21202	0.0 %	21202				
7.5	7.500	7.411	-0.089	30727	0.0 %	30727				
10	10.000	10.016	0.016	41485	0.0 %	41485				

LB93958 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Mar 2018 14:43:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7471B	PPB	16 Mar 2018 14:48:09	S	Std
0.2 - 1	50.2	966	-	0.2000	7471B	PPB	16 Mar 2018 14:50:26	S	Std
2.5 - 1	52.5	10394	-	2.5000	7471B	PPB	16 Mar 2018 14:52:44	S	Std
5.0 - 1	55	21202	-	5.0000	7471B	PPB	16 Mar 2018 14:55:01	S	Std
7.5 - 1	57.5	30727	-	7.5000	7471B	PPB	16 Mar 2018 14:57:18	S	Std
10 - 1	510	41485	-	10.0000	7471B	PPB	16 Mar 2018 14:59:36	S	Std
ICV50 - 1	ICV50	15373	3.6921	-	7471B	PPB	16 Mar 2018 15:02:22	U	SMPL
ICB50 - 1	ICB50	44	-0.0205	-	7471B	PPB	16 Mar 2018 15:04:38	U	SMPL
CCV89 - 1	CCV89	21215	5.1070	-	7471B	PPB	16 Mar 2018 15:06:54	U	SMPL
CCB89 - 1	CCB89	17	-0.0271	-	7471B	PPB	16 Mar 2018 15:09:10	U	SMPL
CRA - 1	CRA	951	0.1991	-	7471B	PPB	16 Mar 2018 15:11:27	U	SMPL
HighStd - 1	HighStd	42270	10.2065	-	7471B	PPB	16 Mar 2018 15:13:42	U	SMPL
ChkStd - 1	ChkStd	29092	7.0148	-	7471B	PPB	16 Mar 2018 15:15:59	U	SMPL
PB107408BL - 1	PBS	-179	-0.0745	-	7471B	PPB	16 Mar 2018 15:18:17	U	SMPL
PB107408BS - 1	LCSS	15470	3.7156	-	7471B	PPB	16 Mar 2018 15:20:35	U	SMPL
J1750-01 - 1	HR-05-031518-A	4092	0.9599	-	7471B	PPB	16 Mar 2018 15:22:53	U	SMPL
J1750-03 - 1	HR-05-031518-B	9173	2.1905	-	7471B	PPB	16 Mar 2018 15:25:10	U	SMPL
J1750-05 - 1	HR-05-031518-C	2595	0.5973	-	7471B	PPB	16 Mar 2018 15:27:28	U	SMPL
J1756-05 - 1	OR-02-031518	4444	1.0451	-	7471B	PPB	16 Mar 2018 15:29:46	U	SMPL
J1938-01 - 1	CS-1	3031	0.7029	-	7471B	PPB	16 Mar 2018 15:32:02	U	SMPL
CCV90 - 1	CCV90	21178	5.0981	-	7471B	PPB	16 Mar 2018 15:34:18	U	SMPL
CCB90 - 1	CCB90	39	-0.0217	-	7471B	PPB	16 Mar 2018 15:36:34	U	SMPL
J1938-05 - 1	CS-2	6905	1.6412	-	7471B	PPB	16 Mar 2018 15:38:50	U	SMPL
J1938-09 - 1	CS-3	5132	1.2118	-	7471B	PPB	16 Mar 2018 15:41:05	U	SMPL
J1940-01 - 1	SP-1-5	108903	26.3449	-	7471B	PPB	16 Mar 2018 15:43:21	U	SMPL
J1945-01 - 1	SP-A	1191	0.2573	-	7471B	PPB	16 Mar 2018 15:47:10	U	SMPL
J1945-01DUP - 1	SP-ADUP	1125	0.2413	-	7471B	PPB	16 Mar 2018 15:49:26	U	SMPL
J1945-01MS - 1	SP-AMS	16831	4.0452	-	7471B	PPB	16 Mar 2018 15:51:43	U	SMPL
J1945-01MSD - 1	SP-AMSD	16774	4.0314	-	7471B	PPB	16 Mar 2018 15:54:01	U	SMPL
J1945-01LX5 - 1		112	-0.0041	-	7471B	PPB	16 Mar 2018 15:56:18	U	SMPL
J1945-01A - 1		17726	4.2620	-	7471B	PPB	16 Mar 2018 15:58:36	U	SMPL
J1940-01DLX5 - 1		19138	4.6040	-	7471B	PPB	16 Mar 2018 16:00:52	U	SMPL
CCV91 - 1	CCV91	21145	5.0901	-	7471B	PPB	16 Mar 2018 16:03:08	U	SMPL
CCB91 - 1	CCB91	28	-0.0244	-	7471B	PPB	16 Mar 2018 16:05:24	U	SMPL