



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93961

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/16/2018	09:10
2	J1931-02		1	67.0	152.6	03/16/2018	09:35
3	J1931-02DUP		1	67.0	152.6	03/16/2018	09:55

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP31518

WorkList ID : 109626

Date : 3/15/2018 8:15:24 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/21/2018	✓ Solid	J1931-02	Flash Point	Cool 4 deg C	PSEG02	I61	1465-OIL	03/14/2018	1010A
03/21/2018	Water	J1931-12	Flash Point	Cool 4 deg C	PSEG02	I61	1470-WW	03/14/2018	1010A
03/21/2018	Water	J1931-18	Flash Point	Cool 4 deg C	PSEG02	I61	1472-75-WW	03/14/2018	1010A

Date/Time 3/16/18 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 3/16/18 10:00 AM
 Received by: CP
 Relinquished by: KS