



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93962

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/16/2018	10:30
2	J1931-12		1	>100.0	>212.0	03/16/2018	10:55
3	J1931-12DUP		1	>100.0	>212.0	03/16/2018	11:20
4	J1931-18		1	>100.0	>212.0	03/16/2018	11:45

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP31518

WorkList ID : 109626

Date : 3/15/2018 8:15:24 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/21/2018	Solid	J1931-02	Flash Point	Cool 4 deg C	PSEG02	I61	1465-OIL	03/14/2018	1010A
03/21/2018	Water	J1931-12	Flash Point	Cool 4 deg C	PSEG02	I61	1470-WW	03/14/2018	1010A
03/21/2018	Water	J1931-18	Flash Point	Cool 4 deg C	PSEG02	I61	1472-75-WW	03/14/2018	1010A

Date/Time

Received by:

Relinquished by:

3/16/18 8:30 AM
KS
CP

Date/Time

Received by:

Relinquished by:

3/16/18 11:50 AM
CP
KS

LB 93962