



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93968

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/16/2018	17:25
2	CAL2	1	Water			7.01	03/16/2018	17:26
3	CAL3	1	Water			10.02	03/16/2018	17:27
4	ICV	1	Water			7.00	03/16/2018	17:30
5	CCV1	1	Water			2.01	03/16/2018	17:31
6	J1755-04	1	Solid	20.02	20	8.11	03/16/2018	17:32
7	J1755-04DUP	1	Solid	20.04	20	8.12	03/16/2018	17:33
8	J1963-04	1	Solid	20.05	20	9.64	03/16/2018	17:35
9	CCV2	1	Water			12.02	03/16/2018	17:36

1893968

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1963 WorkList ID : 109695 Date : 3/16/2018 5:14:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2018	Solid	J1755-04	Corrosivity	Cool 4 deg C	PSEG05		OK-01-031618	03/16/2018	9045C
03/16/2018	Solid	J1963-04	Corrosivity	Cool 4 deg C	PSEG01	I62	1-4-RO	03/16/2018	9045C

Date/Time 03-16-18 5:14 PM
 Received by: JP
 Relinquished by: JP

Date/Time 03-16-18 6:00 PM
 Received by: JP
 Relinquished by: JP