



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93972

Temp.: 16.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/16/2018	18:00
2	CAL2	1	Water			7.01	03/16/2018	18:01
3	CAL3	1	Water			10.02	03/16/2018	18:02
4	ICV	1	Water			7.00	03/16/2018	18:03
5	CCV1	1	Water			2.01	03/16/2018	18:05
6	J1960-02	1	Water			11.45	03/16/2018	18:06
7	J1966-08	1	Water			7.15	03/16/2018	18:07
8	J1966-10	1	Water			8.07	03/16/2018	18:11
9	J1966-10DUP	1	Water			8.08	03/16/2018	18:12
10	CCV2	1	Water			12.01	03/16/2018	18:14

10 93972

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1960 WorkList ID : 109696 Date : 3/16/2018 5:15:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2018	Water	J1960-02-17	Corrosivity	Cool 4 deg C	PSEG01	I61	BB	03/16/2018	9040C
03/16/2018	Water	J1966-08-17	Corrosivity	Cool 4 deg C	PSEG01	I41	FRAC1	03/16/2018	9040C
03/16/2018	Water	J1966-10-17	Corrosivity	Cool 4 deg C	PSEG01	I41	FRAC2	03/16/2018	9040C

Date/Time 03-16-18 5:20 PM
Received by: JP
Relinquished by: JP

Date/Time 03-16-18 6:30 PM
Received by: JP
Relinquished by: JP