



TOTAL VOLATILE SOLIDS - SM2540 G

TEMP1 IN: 105 03/19/2018 16:00 TEMP OUT: 105 03/20/2018 08:00
TEMP2 IN: 104 03/20/2018 08:50 TEMP OUT: 105 03/20/2018 09:58
TEMP3 IN: 540 03/20/2018 10:20 TEMP OUT: 560 03/20/2018 11:30
TEMP4 IN: 560 03/20/2018 12:20 TEMP OUT: 560 03/20/2018 13:00

Run Number: LB93990
SUPERVISOR: apatel
ANALYST: Aparna
BalanceID: WC SC-3
OvenID: WC Oven-2

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result %
1	LB93990BL	84.0539	84.0539	100	84.0539	84.0539	84.0539	84.0539	0	0
2	J1900-35	84.7608	84.7608	100	84.7807	84.7807	84.7692	84.7692	0.0115	115

A = Sample Volume (ml)
B = Final Dish + Samplewt Drying @550 (±50) °C (g)
C = Final Dish + SampleWt Drying @103-@105°C (g)
D = Weight (g)
E = Final Empty Dish Weight (g)
F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result mg/L = $\frac{D}{A} \times 100000$