



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB93994

Reagent/Standard	Lot/Log #
conductivity standard	WP62773

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB93994BL		1	20.7	0.85	0.93	03/19/2018	12:54
2	LB93994BS	1413	1	23.4	1360.00	1402.87	03/19/2018	13:05
3	J1900-07		1	21.3	628.00	675.76	03/19/2018	13:15

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$