

7471B


LB 93997

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

42902



A= 0.0000e+000

B= 2.3256e-004

C= 8.0272e-003

Rho= 0.9999723

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.017 | 0.017  | 40    | 0.000      | 40    |       |       |       |       |
| 0.2    | 0.200  | 0.213 | 0.013  | 880   | 0.0 %      | 880   |       |       |       |       |
| 2.5    | 2.500  | 2.452 | -0.048 | 10509 | 0.0 %      | 10509 |       |       |       |       |
| 5.0    | 5.000  | 4.994 | -0.006 | 21440 | 0.0 %      | 21440 |       |       |       |       |
| 7.5    | 7.500  | 7.539 | 0.039  | 32382 | 0.0 %      | 32382 |       |       |       |       |
| 10     | 10.000 | 9.985 | -0.015 | 42902 | 0.0 %      | 42902 |       |       |       |       |

## LB93997 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Mar 2018 14:19:13

| Sample ID        | Extended ID       | μ Abs. | Conc.   | Std Conc     | Method | Units | Date                 | Type | Type |
|------------------|-------------------|--------|---------|--------------|--------|-------|----------------------|------|------|
| 0.0 - 1          | 50                | 40     | -       | 0.00007471B  | PPB    |       | 19 Mar 2018 14:23:00 | S    | Std  |
| 0.2 - 1          | 50.2              | 880    | -       | 0.20007471B  | PPB    |       | 19 Mar 2018 14:25:17 | S    | Std  |
| 2.5 - 1          | 52.5              | 10509  | -       | 2.50007471B  | PPB    |       | 19 Mar 2018 14:27:35 | S    | Std  |
| 5.0 - 1          | 55                | 21440  | -       | 5.00007471B  | PPB    |       | 19 Mar 2018 14:29:52 | S    | Std  |
| 7.5 - 1          | 57.5              | 32382  | -       | 7.50007471B  | PPB    |       | 19 Mar 2018 14:32:09 | S    | Std  |
| 10 - 1           | 510               | 42902  | -       | 10.00007471B | PPB    |       | 19 Mar 2018 14:34:26 | S    | Std  |
| ICV53 - 1        | ICV53             | 16049  | 3.7403  | -7471B       | PPB    |       | 19 Mar 2018 14:37:11 | U    | SMPL |
| ICB53 - 1        | ICB53             | -121   | -0.0201 | -7471B       | PPB    |       | 19 Mar 2018 14:39:26 | U    | SMPL |
| CCV01 - 1        | CCV01             | 21538  | 5.0169  | -7471B       | PPB    |       | 19 Mar 2018 14:41:42 | U    | SMPL |
| CCB01 - 1        | CCB01             | -112   | -0.0180 | -7471B       | PPB    |       | 19 Mar 2018 14:43:58 | U    | SMPL |
| CRA - 1          | CRA               | 873    | 0.2111  | -7471B       | PPB    |       | 19 Mar 2018 14:46:14 | U    | SMPL |
| HighStd - 1      | HighStd           | 43081  | 10.0269 | -7471B       | PPB    |       | 19 Mar 2018 14:48:30 | U    | SMPL |
| ChkStd - 1       | ChkStd            | 29933  | 6.9692  | -7471B       | PPB    |       | 19 Mar 2018 14:50:46 | U    | SMPL |
| PB107440BL - 1   | PBS               | -199   | -0.0383 | -7471B       | PPB    |       | 19 Mar 2018 14:53:04 | U    | SMPL |
| PB107440BS - 1   | LCSS              | 15269  | 3.5590  | -7471B       | PPB    |       | 19 Mar 2018 14:55:22 | U    | SMPL |
| J1755-03 - 1     | OK-01-031618      | 2211   | 0.5222  | -7471B       | PPB    |       | 19 Mar 2018 14:57:39 | U    | SMPL |
| J1946-01 - 1     | SB-104(4-5)       | 15244  | 3.5531  | -7471B       | PPB    |       | 19 Mar 2018 14:59:57 | U    | SMPL |
| J1946-02 - 1     | SB-104(22-23)     | 74870  | 17.4196 | -7471B       | PPB    |       | 19 Mar 2018 15:02:14 | U    | SMPL |
| J1946-04 - 1     | MW-104(4-5)       | 678    | 0.1657  | -7471B       | PPB    |       | 19 Mar 2018 15:04:32 | U    | SMPL |
| J1946-05 - 1     | MW-104(14-15)     | 1549   | 0.3683  | -7471B       | PPB    |       | 19 Mar 2018 15:06:50 | U    | SMPL |
| CCV02 - 1        | CCV02             | 21073  | 4.9087  | -7471B       | PPB    |       | 19 Mar 2018 15:09:06 | U    | SMPL |
| CCB02 - 1        | CCB02             | -109   | -0.0173 | -7471B       | PPB    |       | 19 Mar 2018 15:11:22 | U    | SMPL |
| J1947-01 - 1     | FK-GF-02-013-A    | 1900   | 0.4499  | -7471B       | PPB    |       | 19 Mar 2018 15:13:38 | U    | SMPL |
| J1947-03 - 1     | FK-GF-02-013-B    | 2147   | 0.5073  | -7471B       | PPB    |       | 19 Mar 2018 15:15:53 | U    | SMPL |
| J1947-03DUP - 1  | FK-GF-02-013-BDUP | 2293   | 0.5413  | -7471B       | PPB    |       | 19 Mar 2018 15:18:10 | U    | SMPL |
| J1947-03MS - 1   | FK-GF-02-013-BMS  | 17580  | 4.0964  | -7471B       | PPB    |       | 19 Mar 2018 15:20:26 | U    | SMPL |
| J1947-03MSD - 1  | FK-GF-02-013-BMSD | 19791  | 4.6106  | -7471B       | PPB    |       | 19 Mar 2018 15:22:43 | U    | SMPL |
| J1963-01 - 1     | 1-4-RO            | 1235   | 0.2952  | -7471B       | PPB    |       | 19 Mar 2018 15:25:00 | U    | SMPL |
| J1965-01 - 1     | SB-113(4-5)       | 31385  | 7.3069  | -7471B       | PPB    |       | 19 Mar 2018 15:27:18 | U    | SMPL |
| J1965-02 - 1     | SB-112(6-7)       | 4692   | 1.0992  | -7471B       | PPB    |       | 19 Mar 2018 15:29:35 | U    | SMPL |
| J1947-03LX5 - 1  |                   | 326    | 0.0838  | -7471B       | PPB    |       | 19 Mar 2018 15:31:54 | U    | SMPL |
| J1947-03A - 1    |                   | 18407  | 4.2887  | -7471B       | PPB    |       | 19 Mar 2018 15:34:10 | U    | SMPL |
| CCV03 - 1        | CCV03             | 21107  | 4.9166  | -7471B       | PPB    |       | 19 Mar 2018 15:36:26 | U    | SMPL |
| CCB03 - 1        | CCB03             | -120   | -0.0199 | -7471B       | PPB    |       | 19 Mar 2018 15:38:42 | U    | SMPL |
| J1946-02DLX5 - 1 |                   | 14653  | 3.4157  | -7471B       | PPB    |       | 19 Mar 2018 15:40:59 | U    | SMPL |
| CCV04 - 1        | CCV04             | 21277  | 4.9562  | -7471B       | PPB    |       | 19 Mar 2018 15:43:16 | U    | SMPL |
| CCB04 - 1        | CCB04             | -116   | -0.0189 | -7471B       | PPB    |       | 19 Mar 2018 15:45:32 | U    | SMPL |