



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94004

Temp.: 14.3 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/19/2018	15:00
2	CAL2	1	Water			7.00	03/19/2018	15:01
3	CAL3	1	Water			10.03	03/19/2018	15:02
4	ICV	1	Water			6.99	03/19/2018	15:03
5	CCV1	1	Water			2.01	03/19/2018	15:05
6	J1979-05	1	Water			7.85	03/19/2018	15:06
7	J1979-05DUP	1	Water			7.86	03/19/2018	15:07
8	CCV2	1	Water			12.02	03/19/2018	15:08

LB94004

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1979

WorkList ID : 109752

Date : 3/19/2018 2:41:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2018	Water	J1979-05-65	pH	Cool 4 deg C	PSEG01	J13	A3960	03/19/2018	9040C

Date/Time 03-19-18 2:45 PM
Received by: JP
Relinquished by: CB

Date/Time 03-19-18 3:50 PM
Received by: SP
Relinquished by: JP