



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94010

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/19/2018	09:05
2	J1931-19		1	>100.0	>212.0	03/19/2018	09:35
3	J1931-19DUP		1	>100.0	>212.0	03/19/2018	10:02
4	J1960-02		1	>100.0	>212.0	03/19/2018	10:40
5	J1966-08		1	>100.0	>212.0	03/19/2018	11:15
6	J1966-10		1	>100.0	>212.0	03/19/2018	11:48

Result = (Celsius * 1.8) + 32

LB94010

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP31918

WorkList ID : 109712

Date : 3/19/2018 8:16:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/21/2018	Water	J1931-19	Flash Point	Cool 4 deg C	PSEG02	I61	1482-1484-87-OILY-SNOW	03/14/2018	1010A
03/21/2018	Water	J1960-02	Flash Point	Cool 4 deg C	PSEG01	I61	BB	03/16/2018	1010A
03/21/2018	Water	J1966-08	Flash Point	Cool 4 deg C	PSEG01	I41	FRAC1	03/16/2018	1010A
03/21/2018	Water	J1966-10	Flash Point	Cool 4 deg C	PSEG01	I41	FRAC2	03/16/2018	1010A

Date/Time

Received by:

Relinquished by:

3/19/18 8:30 AM
KS
CP

Date/Time

Received by:

Relinquished by:

3/19/18 12:00 PM
CP
KS