



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB94012

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/19/2018	16:45
2	CAL2	1	Water			7.00	03/19/2018	16:46
3	CAL3	1	Water			10.03	03/19/2018	16:47
4	ICV	1	Water			7.01	03/19/2018	16:48
5	CCV1	1	Water			2.01	03/19/2018	16:50
6	J1760-04	1	Solid	20.02	20	8.90	03/19/2018	16:51
7	J1974-02	1	Solid	20.04	20	8.34	03/19/2018	17:00
8	J1975-07	1	Solid	20.05	20	6.45	03/19/2018	17:01
9	J1977-04	1	Solid	20.02	20	8.73	03/19/2018	17:02
10	J1979-04	1	Solid	20.03	20	8.51	03/19/2018	17:03
11	J1979-04DUP	1	Solid	20.05	20	8.52	03/19/2018	17:05
12	CCV2	1	Water			12.02	03/19/2018	17:07

LB 94012

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1975 WorkList ID : 109748 Date : 3/19/2018 2:05:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2018	Solid	J1760-04	Corrosivity	Cool 4 deg C	PSEG05	I62	SU-03-031918	03/19/2018	9045C
03/19/2018	Solid	J1974-02	Corrosivity	Cool 4 deg C	PSEG01	I62	YR18020063	03/19/2018	9045C
03/19/2018	Solid	J1975-07	Corrosivity	Cool 4 deg C	PSEG01	I61	FS-SOIL	03/19/2018	9045C
03/19/2018	Solid	J1977-04	Corrosivity	Cool 4 deg C	PSEG01	J11	SP	03/19/2018	9045C
03/19/2018	Solid	J1979-04	Corrosivity	Cool 4 deg C	PSEG01	J13	SP-1	03/19/2018	9045C

Date/Time 03-19-18 2:10 PM
Received by: SP
Relinquished by: SP

Date/Time 03-19-18 5:35 PM
Received by: SP
Relinquished by: SP